



በኢትዮጵያ ፌዴራላዊ ዲሞክራሲያዊ ሪፐብሊክ
THE FEDERAL DEMOCRATIC REPUBLIC OF ETHIOPIA
 የመድኃኒት ፈንድና አቅርቦት ኤጀንሲ
PHARMACEUTICALS FUND AND SUPPLY AGENCY



#T/C PFSA/06/495/04
 Ref.No.

ቀን 17/8/2017
 Date

TO: All participants of Tender No ICB/PFSA6/MOH-SDG/MI/45/16

Dear Sirs,


Ref: - Technical Evaluation result

Assessment of above captioned tender is now finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceuticals Fund and Supply Agency (PFSA) Director General Office from August 17,2017 to August 23, 2017; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards

Bekel Ashagre
 Pharmaceuticals & Medical
 Supplies Procurement
 Directorate Director



ENC: 2 page
 CC

- Director General
- Operation Deputy Director General
- MIS Directorate
- Medical Device & Supplier Procurement Team

PFSA

ስልክ 011-275 17 70 }
 ገጽ 011-275 06 74 } ማዘሪያ
 011- 276 06 76
 011-276 32 65 ዋና ዳይሬክተር

011-276 32 73
 011-276 32 66
 011-276 32 72 የመድኃኒትና የሕክምና መገልገያዎች
 ክምችት ሥርዓት ዳይሬክቶሬት
 011-276-32 68 - ፈንድ አስተዳደር ዳይሬክቶሬት

ፋክስ 011-275 25 55 ፖ.ሣ.ቁ
 ፊክስ: 011-155 32 77 P.O.Box: 21904
 E-mail: pfsa@ethionet.et
 www.pfsa.gov.et

እባክዎ መልስ ሲሰጡ የኛን ቁጥር ይጥቀሱ

Technical Evaluation Result

ICB/PFSA6/MOH-SDG/MI/45/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason of Non Compliance
1	Hand Held Dynamometer	Each	40	BEARDELL	NOT COMPLY	DATA STORAGE CAPACITY AND TIME TO REACH PEAK NOT INDICATED
2	Doppler [For Arterial And Venous Examination]	Each	40	BEARDELL	COMPLY	NO BIDDER
3	Inclinometer	Each	80	BEARDELL	COMPLY	NO BIDDER
4	Goniometers	Pack	20	BEARDELL	COMPLY	NO BIDDER
5	Spirometer	Each	80	BEARDELL	COMPLY	NO BIDDER
6	Quadriceps Bench	Each	60	BEARDELL	NOT COMPLY	GRAPHIC CURVE DISPLAY, STORAGE MEMMORY AND CLINICAL INTERPRETATION NOT INDICATED
				BOSTON MEDICAL FZE	COMPLY	
7	Body Weight Support Treadmill	Each	40	ENRAF NONIOUS	COMPLY	
				BEARDELL	NOT COMPLY	NO FMHACA Product Certificate
8	Continuous Passive Motion Machine	Each	40	ENRAF NONIOUS	NOT COMPLY	NO WIRELESS HEART RATE MONITOR
9	Electromyogram Machine (EMG)	Each	40	NIHON KOHDEN	COMPLY	NO BIDDER
10	Muscle Stimulator	Each	40	ENRAF NONIOUS	COMPLY	
11	Therapeutic Ultrasound	Each	60	ENRAF NONIOUS	COMPLY	
				BEARDELL	NOT COMPLY	NO FMHACA Product Certificate
12	Diathermy	Each	40	ENRAF NONIOUS	COMPLY	
				BEARDELL	NOT COMPLY	NO FMHACA Product Certificate
13	Iontophoresis Electrode	Pack	100	ENRAF NONIOUS	COMPLY	
				BEARDELL	NOT COMPLY	NO FMHACA Product Certificate
14	Postural Training Mirror	Each	40	BEARDELL	COMPLY	
				BEARDELL	COMPLY	

[Signature]

[Signature]

Technical Evaluation Result

ICB/PFSA6/MOH-SDG/MI/45/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason of Non Compliance
15	Parallel Walking Bar	Each	20	ENRAF NONIOUS BEARDESELL	COMPLY NOT COMPLY	NO FMHACA Product Certificate
16	Ultrasound Machine [For Obstetric Scanning]	Each	900	EMPEROR MEDICAL MINDRAY BOSTON MEDICAL FZE CAREWELL SONOSCAPE MEDICAL ELSMED HEALTH CARE SOLI SHANTOU	NOT COMPLY COMPLY COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY	12 inch display against ≥ 15 inch NO FMHACA Product Certificate 12 inch display against ≥ 15 inch NO FMHACA Product Certificate
17	Ultrasound General Purpose	Each	350	EMPEROR MEDICAL CAREWELL MINDRAY BOSTON MEDICAL FZE SONOSCAPE MEDICAL SAMSUNG ELSMED HEALTH CARE SOLI SHANTOU	NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY	list of spare part with their part number not indicated Linear probe 5-10MHz Against 4-10MHz Linear probe 5-10MHz Against 4-10MHz Linear probe 5-10MHz Against 4-10MHz Model Difference between Compliance Sheet and Attached Catalogue Linear probe 5-12MHz Against 4-10MHz Linear probe 5-12MHz Against 4-10MHz Linear probe 5-13MHz Against 4-10MHz
18	Color Doppler Ultrasound Machine	Each	44	MINDRAY BOSTON MEDICAL FZE SONOSCAPE MEDICAL ELSMED HEALTH CARE SOLI SHANTOU	NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY	list of spare part with their part number not indicated RESPIRATION TRIGGER FACILITY NOT INDICATED TEE Probe Not indicated RESPIRATION TRIGGER FACILITY NOT INDICATED RESPIRATION TRIGGER FACILITY NOT INDICATED TEE Probe Not indicated
19	Stationary Bicycle	Each	60	SHANTOU	NOT COMPLY	RESPIRATION TRIGGER FACILITY NOT INDICATED TEE Probe Not indicated

End

NO BIDDER

