



በኢትዮጵያ ፌዴራላዊ ዲሞክራሲያዊ ሪፐብሊክ
THE FEDERAL DEMOCRATIC REPUBLIC OF ETHIOPIA
 የመድኃኒት ፈንድና አቅርቦት ኤጀንሲ
PHARMACEUTICALS FUND AND SUPPLY AGENCY



ቁጥር PFSA/06/1494/04
 Ref.No.

ቀን 17/18/2017
 Date

TO: All participants of Tender No ICB/PFSA6/RDF/MI/36/16

Dear Sirs,

Ref: - Technical Evaluation result

Assessment of above captioned tender is now finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceuticals Fund and Supply Agency (PFSA) Director General Office from August 17,2017 to August 23, 2017; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards,

Bekete Abay
 Bekete Abay
 Pharmaceuticals & Medical
 Supplies Procurement
 Directorate Director



ENC: 10 page
 CC

- Director General
- Operation Deputy Director General
- MIS Directorate
- Medical Device & Supplier Procurement Team

PFSA

ስልክ 011-275 17 70 }
 Tel. 011-275 06 74 } ማዘሪያ
 011- 276 06 76
 011-276 32 65 ዋና ዳይሬክተር

011-276 32 73
 011-276 32 66
 011-276 32 72 የመድኃኒትና የሕክምና መገልገያዎች
 ክምችት ሥርዓት ዳይሬክቶሬት
 011-276-32 68 - ፈንድ አስተዳደር ዳይሬክቶሬት

ፋክስ 011-275 25 55 ፖ.ሣ.ቁ 21904
 Fax: 011-155 32 77 P.O.Box:
 E-mail: pfsa@ethionet.et
 www.pfsa.gov.et

እባክዎ መልስ ሲሰጡ የኛን ቁጥር ይጥቀሱ

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
1	Anesthesia Machine	Each	4	Boston Medical FZE ✓ Suzhou Industrial Quayle Dental Narula Beijing Aeonmed Acoma Mindray ✓ Comen	COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY COMPLY NOT COMPLY	No FMHACA Product Certificate Patient Monitor Not Indicated Catalogue Not Original Model & Catalogue Not Indicated For Patient Monitor O2 & N2O Cylinder Not Quoted Two Flow Meters Against Three Flow Meters
2	Laundry Machine	Each	17	Narula Omedad ✓ Alliance ✓ Tolon Italclean Avishkar Beardsell Suzhou Industrial	NOT COMPLY COMPLY COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY	Catalogue Not Original Washer Extractor Capacity 40kg Against 50kg Per Cycle Dryer 36kg Against 40kg Catalogue Not Original List Of Spare Part With Their Part Number Is Not Attached Accessories Not Indicated
3	Autoclave 75L, Vertical Type	Each	4	Apothecaries Beardsell	NOT COMPLY NOT COMPLY	List of Spare Part With Their Part Number Is Not Attached No Water Softener
4	Automatic External Defibrillator	Each	1	Mindray ✓ Beardsell Narula	COMPLY NOT COMPLY NOT COMPLY	Catalogue Is Not Supported With Detail Specification Catalogue Not Original
5	Capnography	Each	2	EDAN ✓ Mindray	COMPLY COMPLY	
6	Cardiotocography Monitor	Each	2	Boston Medical FZE EDAN Apothecaries	COMPLY COMPLY NOT COMPLY	Catalogue Is Not Supported With Detail Specification & No Spare Part List With Part Their Number
7	CD4 Machine	Each	2	BD ✓ Mindray	COMPLY NOT COMPLY	Coring Station Not Indicated, Pre Programmed Electronic Pipette Not Quoted.
8	Centrifuge	Each	15	Quayle Dental Fanem Infinity Medlab Apothecaries Avishkar Beardsell	NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached 5000 RPM Against 6000 RPM Catalogue Not Original Swing Out Rotor Against Fixed Angle Rotor, List Of Spare Part With Their Part Number Is Not Attached Catalogue Not Original Speed 5500 RPM Against 6000RPM

[Handwritten Signature]

[Handwritten Signature]

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
				Narula	NOT COMPLY	Catalogue Not Original
				Suzhou Industrial ✓	COMPLY	
				HEMC	NOT COMPLY	Automatic Break and Lid Interlock Not Indicated
				Human	NOT COMPLY	RPM 5000 Against 6000 & No List Of Spare Part With Their Part Number
9	Centrifuge, Hematocrite	Each	10	HEMC	NOT COMPLY	Automatic Break and Lid Interlock Not Indicated
				Quayle Dental	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Infinity Medlab ✓	NOT COMPLY	Catalogue Not Original
				Nuve ✓	COMPLY	
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, Speed RPM 11500 Against 12000
				Avishkar	NOT COMPLY	Catalogue Not Original
				Fanem	NOT COMPLY	Adaptor For Smaller Tubes Not Indicated
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Narula	NOT COMPLY	Catalogue Not Original
				Beardsell	NOT COMPLY	No FMHACA Product Certificate
				Suzhou Industrial	NOT COMPLY	200 Capillary Tube Is Not Indicated
10	Coagulation Analyzer	Each	2	Diagnostic Stago ✓	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Suzhou Industrial	NOT COMPLY	20 Sample Position Against 50 , 40 Sample/Hr Against 260
				SYSMEX	NOT COMPLY	Throughput 180 Tests Per Hour Against 260 Tests Per Hour
				Human	NOT COMPLY	Sample Position 25 Against 50
11	Flexible Colonoscope	Set	1	No bidder		
12	Echocardiography with Cardiac Probe	Each	1	Boston Medical FZE	COMPLY	
				SonoScape ✓	COMPLY	
				Mindray	COMPLY	
13	Dental Unit	Each	1	Quayle Dental ✓	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
14	Color Doppler Ultrasound Machine	Each	2	Boston Medical FZE	NOT COMPLY	TEE Probe Not Indicated
				SonoScape	NOT COMPLY	Respiration Trigger Not Indicated
				Mindray	NOT COMPLY	Respiration Trigger Not Indicated
15	Fully Digital Radiography Machine (with Out Fluoroscopy)	Each	1	Boston Medical FZE ✓	COMPLY	
				Wangdong	NOT COMPLY	Tube Current 25-630mAmp Against 10-8000mAmp
				Suzhou Industrial	NOT COMPLY	Tube Current 10-640mAmp Against 10-800mAmp
16	Electrical Centrifuge ,10000 RPM	Each	2	Quayle Dental	NOT COMPLY	Rotor 12x15ml Against 12x20ml
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached

(Handwritten signatures and initials)

(Handwritten signature)

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
				Avishkar	NOT COMPLY	Catalogue Not Original
				Fanem	NOT COMPLY	RPM 5000 Against 10000
				Beardsell ✓	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
				Human	NOT COMPLY	RPM 5000 Against 10000
17	Electrical Centrifuge, 24 Tube	Each	2	Quayle Dental	NOT COMPLY	12 Tube Against 24 Tube
				Avishkar	NOT COMPLY	Catalogue Not Original
				Fanem	NOT COMPLY	No Product Certificate
				Beardsell	NOT COMPLY	18 Tube Against 24 Tube
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
18	Electrical Centrifuge, 36 Tube	Each	2	Quayle Dental	NOT COMPLY	12 Tube Against 36 Tube
				Fanem	NOT COMPLY	No Product Certificate
				Avishkar	NOT COMPLY	Catalogue Not Original
				Appothecaries	NOT COMPLY	Catalogue Not Attached
				Beardsell	NOT COMPLY	18 Tube Against 36 Tube
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
19	ECG, Six Channel	Each	1	SonoScape	COMPLY	
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, Paper Size Not A4, Paper Speed Not Indicated In Catalogue
				Boston Medical FZE	COMPLY	
				EDAN	COMPLY	
				Suzhou Industrial	NOT COMPLY	Accessories Not Indicated
20	Electrolyte Analyzer	Each	2	Suzhou Industrial	NOT COMPLY	Electrode Mg++, Li- Not Indicated, Built In Thermal Printer and Analyze Time Not Indicated
				Human	NOT COMPLY	Electrode Mg++, Li- Not Indicated
21	Electrosurgical Unit	Each	15	HEMC	NOT COMPLY	Cutting Mode Is Not Indicated (Soft, Force,...Bipolar)
				Quayle Dental	NOT COMPLY	Coagulation 150w Against 200w
				Boston Medical FZE	COMPLY	
				Acoma	NOT COMPLY	Accessories not Fully Quoted
				Avishkar	NOT COMPLY	Catalogue Not Original
				Bowa ✓	COMPLY	
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, Foot Switch Not Indicated
				Narula	NOT COMPLY	Catalogue Not Original
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Medtronic	NOT COMPLY	Mono Foot Switch Against Two Pedal Switch
22	Endoscope	Each	1	Boston Medical FZE	NOT COMPLY	Laparoscopy Not Indicated
				Medtronic	NOT COMPLY	Colonoscopy Not Indicated

[Handwritten Signature]

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non-Compliance
23	ESR Analyzer	Each	2	Avishkar	NOT COMPLY	Catalogue Not Original
				Boston Medical FZE	COMPLY	
				Suzhou Industrial	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
24	Glucometer	Each	17	Narula	NOT COMPLY	Catalogue Not Original
25	Hematology Analyzer 5 Differential	Each	3	Mindray	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Abbot	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Nihon Kohden ✓	COMPLY	
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Boston Medical FZE	COMPLY	
				SYSTEMEX	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
26	Hematology Analyzer 3 Differential	Each	5	Meril Diagnostic	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, Installation and After Sales Services Are Not Indicated
				Mindray	COMPLY	
				Abbot ✓	COMPLY	
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Nihon Kohden	COMPLY	
				Boston Medical FZE	COMPLY	
				Suzhou Industrial	NOT COMPLY	No Product Certificate
				Meril Diagnostic	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, Installation and After Sales Services Are Not Indicated
				SYSTEMEX	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
27	Hormone Analyzer	Each	2	No Bidder		
28	Infant Radiant Warmer	Each	11			
				HEMC	NOT COMPLY	Over Head Exam Light Not Indicated, Preset & Actual Skin, Air Elapsed/Remained Temperature Not Indicated
				Quayle Dental	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Avishkar	NOT COMPLY	Catalogue Not Original
				Fanem ✓	COMPLY	
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Ningbo David ✓	COMPLY	
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
				Boston Medical FZE	NOT COMPLY	No Integrated Timer
29	Infusion Pump	Each	3	Avishkar	NOT COMPLY	Catalogue Not Original
				Boston Medical FZE ✓	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, Battery Life 3 Hr Against 8 Hr

[Signature]

[Signature]

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
30	Lensometer	Each	5	Suzhou Industrial	NOT COMPLY	Catalogue Not Attached
31	Medical Refrigerator Large Volume 650L	Each	8	Avishkar	NOT COMPLY	Catalogue Not Original
				Quayle Dental ✓	COMPLY	
				ABS International FZE	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Avishkar	NOT COMPLY	Catalogue Not Original
				B Medical ✓	COMPLY	
				Beardsell	NOT COMPLY	Power Consumption 900W Against 320W
				Boston Medical FZE	COMPLY	
32	Mechanical Ventilator	Each	10	Suzhou Industrial	NOT COMPLY	Power Consumption 443W Against 320W
				Mindray	COMPLY	
				Quayle Dental	NOT COMPLY	Catalogue Is Not Supported With Detail Specification
				Boston Medical FZE	COMPLY	
33	Patient Monitor	Each	7	Suzhou Industrial	NOT COMPLY	Catalogue Is Not Supported With Detail Specification
				Suzhou Industrial	NOT COMPLY	Separated CPU Module Is Not Supported
				Mindray	NOT COMPLY	Separated CPU Module Is Not Supported
				Boston Medical FZE	NOT COMPLY	Separated CPU Module Is Not Supported
				Comen ✓	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
				EDAN	COMPLY	
34	O.R Light, Mobile	Each	24	HEMC	NOT COMPLY	Integrated Battery Is Not Indicated
				Quayle Dental	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Avishkar	NOT COMPLY	Catalogue Not Original
				Appothecaries	NOT COMPLY	Battery Capacity Not Indicated
				Dr. Mach ✓	COMPLY	
				Beardsell	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Boston Medical FZE	COMPLY	
				Suzhou Industrial	NOT COMPLY	Built-In Battery Not Indicated
				Mindray	COMPLY	
				Infinity Mediab	NOT COMPLY	Catalogue Not Original
				Beijing Aeonmed ✓	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
35	Mobile X-Ray System, Digital	Each	2	Boston Medical FZE	COMPLY	
				Wangdong	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				SHIMADZU	NOT COMPLY	AEC Not Supported
36	Microscope, Binocular	Each	22	Suzhou Industrial	NOT COMPLY	Accessories and Consumables Are Not Indicated

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
				HEMC	NOT COMPLY	Mechanical Stage 120x120mm Against 200x250mm, No Interpupillary Distance and Diopter Adjustment
				Avishkar	NOT COMPLY	Catalogue Not Original
				Apothecaries	NOT COMPLY	Mechanical Stage and Travelling Distance, Coarse and Fine Focusing, Eye Guard Retractable Not Indicated In Catalogue
				Beijing Teda	NOT COMPLY	Retractable Eye Guard Not Indicated
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Beardsell	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
37	Semen Analyzer	Each	1	No Bidder		
38	Generator Set	Each	1			
				Beardsell	COMPLY	Provisionally Accepted
				Suzhou Industrial	COMPLY	
39	Steam Sterilizer 25 L	Each	36			
				Apothecaries	NOT COMPLY	Power Consumption 1.85kw Against 1.5kw, List Of Spare Part With Their Part Number Is Not Attached
				Nuve	COMPLY	
				Beardsell	NOT COMPLY	No FMHACA Product Certificate, No List Of Spare Part With Their Part Number
				Boston Medical FZE	COMPLY	
40	Contact Tonometer	Each	5	Narula	NOT COMPLY	Catalogue Not Original
41	Water Bath 8 L with Lid	Each	7	No Bidder		
				Fanem	NOT COMPLY	Quoted Model Not Indicated In Catalogue
				Avishkar	NOT COMPLY	Catalogue Not Original
				Karl hecht	NOT COMPLY	Bath Volume Not Indicated
				Quayle Dental	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
42	Water Distiller 20L/Hr, with Tank	Each	7			
				Quayle Dental	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
43	Vacuum Extractor Electrical	Each	105			
				HEMC	NOT COMPLY	Modified Bird Cup 60mm Widest, 50cm At Opening and Flexion Point At 6 & 11 Not Indicated
				Apothecaries	NOT COMPLY	Quoted Item Is Not Vacuum Extractor, No Integral Palm Pump
				Avishkar	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
44	Blood Bank Refrigerator	Each	15			
				Fanem	NOT COMPLY	Electronic Fuse Auto Disconnect/Connect Not Indicated
				Beardsell	NOT COMPLY	Power Consumption 460W Against 250W
				B Medical	COMPLY	
				ABS International FZE	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Nuve	COMPLY	
				Quayle Dental	NOT COMPLY	Volume 388L Against 450L




Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
				Boston Medical FZE	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Suzhou Industrial	NOT COMPLY	Automatic Voltage Regulator Not Indicated. Ambient Temperature 10-32 Deg C Against 10-43 Deg C
45	General Purpose Ultrasound	Each	17		COMPLY	
				Mindray	COMPLY	
				SonoScape	COMPLY	
				Boston Medical FZE	COMPLY	
				EDAN	NOT COMPLY	No Detail Specification, 2 - 6 MHz Against 5 - 7.5 MHz, 6-10MHz Against 5-7.5mhz
				Quayle Dental	NOT COMPLY	Beam former Capacity 3.5 MHz to 7.5 MHz Against 3.0 MHz to 9.0 MHz
46	Chemistry Machine	Each	1		NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Human	NOT COMPLY	Catalogue Not Original
				Avishkar	NOT COMPLY	Catalogue Not Original
				Mindray	COMPLY	
				Suzhou Industrial	NOT COMPLY	Catalogue is not supported with detail specification
47	Surgical Suction Machine, Elec, 2 Bottle	Each	30		NOT COMPLY	Flow Meter 35L/min Against 60L/min
				HEMC	NOT COMPLY	
				Appothecaries	NOT COMPLY	Vacuum Capacity 600mmHg Against 900mmHg, List Of Spare Part With Their Part Number Is Not Attached
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Avishkar	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
48	Oxygen Concentrator	Each	60		NOT COMPLY	Oxygen Concentration: 95% ±3% At 1-3 LPM, 92% ±3% At 4 LPM, 90% ±3% At 5 LPM Is Not Indicated
				HEMC	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Appothecaries	NOT COMPLY	Catalogue Not Original
				Avishkar	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Longfian	NOT COMPLY	Catalogue Not Original
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Quayle Dental	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
49	Pulse Oximeter	Each	50		NOT COMPLY	High and Low Alarm Setting & Oxygen Concentration Range Is Not Indicated
				HEMC	NOT COMPLY	Oxygen Saturation and Pulse Rate Measurement Range Is Not Indicated
				Appothecaries	COMPLY	Catalogue Not Original
				Avishkar	NOT COMPLY	Catalogue Not Original
				Beijing Teda	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Medtronic	COMPLY	
50	Deep Freezer with Thermometer 350Lit	Each	4		NOT COMPLY	Power Consumption 860W Against 500W
				Beardsell	NOT COMPLY	Catalogue Not Original
				Avishkar	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				ABS International FZE	NOT COMPLY	
				B Medical	COMPLY	
				Quayle Dental	NOT COMPLY	Ambient Temperature Not Indicated

(Handwritten signatures and initials)

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
51	Refrigerator 250lit		24	Suzhou Industrial	NOT COMPLY	Automatic Voltage Regulator Not Indicated. Ambient Temperature 10-32 Deg C Against 10-43 Deg C
				Suzhou Industrial	NOT COMPLY	Automatic Voltage Regulator Not Indicated. Ambient Temperature 10-32 Deg C Against 10-43 Deg C
				Avishkar	NOT COMPLY	Catalogue Not Original
				Fanem	NOT COMPLY	No Auto Electric Disconnect/Connect Electronic Fuse
				B Medical	COMPLY	
				ABS International FZE	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Beardsell	COMPLY	
				Quayle Dental	NOT COMPLY	Ambient Temperature Not Indicated
				No Bidder		
				HEMC	NOT COMPLY	Operating Time Not Indicated
				Avishkar	NOT COMPLY	Catalogue Not Original
				Beardsell	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached, One Tray Against Three Tray
				Infinity Medlab	NOT COMPLY	Catalogue Not Original
				Appothecaries	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Nuue	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
54	Vital Sign Monitor	Each	12			
				Mindray	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Boston Medical FZE	COMPLY	
				Appothecaries	NOT COMPLY	NIBP Measurement Is Not Indicated
				Narula	NOT COMPLY	Catalogue Not Original
				Suzhou Industrial	NOT COMPLY	Catalogue Is Not Supported With Detail Specification
				EDAN	NOT COMPLY	Respiration Rate 0 -150 Bpm Against 6 -180 Bpm
				Comen	COMPLY	
55	Analytical Balance	Each	9			
				Beardsell	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Narula	NOT COMPLY	Catalogue Not Original
56	Shaker Rotary Blood Specimen	Each	10			
				Quayle Dental	COMPLY	
				Narula	NOT COMPLY	Catalogue Not Original
				Avishkar	NOT COMPLY	Catalogue Not Original
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Fanem	NOT COMPLY	Quoted Model Not Indicated In Catalogue
				Karl Hecht	NOT COMPLY	16 X 8 ml Against 20 X 15 ml Test Tube
57	Rotary Agglutination Test	Each	16			
				Fanem	NOT COMPLY	Quoted Model Not Indicated In Catalogue
				Avishkar	NOT COMPLY	Catalogue Not Original
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached & RPM 80 Against 100
				Quayle Dental	COMPLY	

Technical Result

ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
58	Vortex, Test Tube	Each	8	Narula	NOT COMPLY	Catalogue Not Original
				Quayle Dental	COMPLY	
				Karl Hechet	COMPLY	
				Avishkar	NOT COMPLY	Catalogue Not Original
				Human	NOT COMPLY	List Of Spare Part With Their Part Number Is Not Attached
				Fanem	NOT COMPLY	Quoted Model Not Indicated In Catalogue

End of Item Based Result

1	LOT I: Instrument sets (Item 1- 88)					
	Quayle Dental				COMPLY	
	Avishkar				NOT COMPLY	Sample is not submitted
	Boston Medical FZE				NOT COMPLY	Incomplete lot, sample is not submitted
	Appothecaries				COMPLY	
	Trudent				NOT COMPLY	Sample is not submitted
	Deluxe				NOT COMPLY	Incomplete lot, sample is not submitted
	Kay and Company				NOT COMPLY	Sample is not submitted
2	LOT II: Laboratory glass wares and instruments (Item 1- 69)					
	Deluxe				NOT COMPLY	Incomplete lot, Catalogue not original
	ABS international				NOT COMPLY	Incomplete lot
	Quayle Dental				NOT COMPLY	No compliance sheet with specification, original manufacturer catalogue is not attached
	Avishkar				NOT COMPLY	Catalogue not original
	Boston Medical FZE				NOT COMPLY	Incomplete lot
	Appothecaries				NOT COMPLY	Catalogue not original
	Karl Hecht				NOT COMPLY	Incomplete lot
3	LOT III: Hospital furnitures (Item 1- 26)					
	Quayle Dental				NOT COMPLY	Original manufacturer catalogue is not attached for quoted item no 3 & 10
	Boston Medical FZE				NOT COMPLY	Incomplete lot
	Deluxe				NOT COMPLY	Incomplete lot, Catalogue not original
	Appothecaries				NOT COMPLY	Carrying capacity and frame diameter not indicated for item no 5, 8. lifting height, overall size is not indicated for item no 10
	Avishkar				NOT COMPLY	Catalogue not original, copy paste of PFSA specification
	Suzhou Industrial				NOT COMPLY	Incomplete lot
	Beardsell				NOT COMPLY	Different Model in Compliance Sheet and Cattelogue for item 4,11 and 15; Item11 Mortuary trolley height adjustment 775-1140mm against 400-1725mm
	Linet				NOT COMPLY	Incomplete lot
	Trigenesis				NOT COMPLY	Catalogue not original, 200g against 100g for item no 12, 625x475x920mm against 860x560x1040mm for item no 9

[Signature]

[Signature]

[Signature]

Technical Result
ICB/PFSA6/RDF/MI/36/16

S.No	Item	Unit	Qty	Bidder	Compliance	Reason Of Non Compliance
				Kay and Company	NOT COMPLY	Foam density is not indicated, 1830x600x700mm against 1800x800x750mm for Gyn.Ex. coach, and list of spare part for patient bed
				Narula	NOT COMPLY	Catalogue not original and model not indicated
				Infinity Medlab	NOT COMPLY	Catalogue not original and model not indicated
				HEMC	NOT COMPLY	Weighing adult scale measuring range and graduation not indicated for item no. 12 and baby scale graduation not indicated for item no. 17
4	LOT IV: Diagnostic Instruments (Item 1- 21)					
				Apothecaries	NOT COMPLY	Original manufacturer catalogue is not attached for item no 4,5,15,17 and 20
				Narula	NOT COMPLY	Catalogue not original and model is not indicated
				Karl Hecht	NOT COMPLY	Incomplete lot
				Trudent	NOT COMPLY	Incomplete lot
				Deluxe	NOT COMPLY	Incomplete lot
				Boston Medical FZE	NOT COMPLY	Incomplete lot
				Avishikar	NOT COMPLY	Catalogue not original, Copy paste of PFSA specification.
				Kay and Company	COMPLY	
				Infinity Medlab	NOT COMPLY	Catalogue not original, model not indicated
				HEMC	NOT COMPLY	Quoted item for item no. 17 is not digital hemoglobinometer
5	LOT V: Miscellaneous instruments (Item 1- 24)					
				Deluxe	NOT COMPLY	Catalogue not original
				Apothecaries	NOT COMPLY	Original manufacturing catalogue is not attached for item no 9,10,11
				Avishikar	NOT COMPLY	Catalogue not original

End of Lot Based Result