

ቀን 19 JUN 2020
Date

ቁጥር
Ref. No. EPISA/06/1119/04

TO: All Participants of Tender no. ICB/PSA6/COVID-19/MOH/MI/08/20

Dear Sirs,

Ref: - Tender Result

Assessment of the above captioned tender is now finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceutical Supply Agency (PSA) Director General Office from June 19, 2020, to June 25, 2020; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

Moreover, bidders with 2nd least price kindly requested to submit your consent confirmation letter till 25 June, 2020 to supply the allocated quantity with by first winner price. The one who fails to do so could be considered as unwilling to accept the allocated Quantity.

With Best Regards,


ENC: 4 page



CC

- Director General
- Deputy Director General /Procurement/
- Medical Equipment Procurement Team

PSA



Company Name:		Document No: OF/TM/ HO-16	
የኢትዮጵያ መድሃኒት አቅርቦት ኤጀንሲ			
Ethiopian Pharmaceuticals Supply Agency			
Title:		Effective date :	
Official Partial Winner list		Rev. No.: 0	Page 1 of 1

Tender No.: ICB/PSA6/COVID-19/MOH/MI/08/20

Ser No.	Description	Unit	Quantity	Proposed Winner list	Currency	Incoterm	Unit Price		Installation cost in ETB	Payment Term	Delivery Time	Remark
							Foreign	Birr				
1	Ventilator - ICU, Mechanical	Each	400	1st winner: Shenzhen Mindray & 2nd winner Boston Medical FZE	USD	FOB	6,993.00	241070.388	3,916.00	LC	as per the bid	
	Disposable breathing circuit (adult)				USD	FOB	10.00	344.731				
	Disposable breathing circuit (child)				USD	FOB	9.00	310.2579				
	NIV mask(large, inc. headgear)				USD	FOB	80.00	2757.848				
	NIV mask(medium, inc. headgear)				USD	FOB	80.00	2757.848				
	oxygen therapy large				USD	FOB	30.00	1034.193				
	Nebulizer accessories kit for ventilation				USD	FOB	10.00	344.731				
	Bacteria filter				USD	FOB	2.00	68.9462				
	Viral filter				USD	FOB	2.00	68.9462				
	2	Pulse Oximeter	Each		6665	1st winner:Shenzhen Creative & 2nd winner Guangdong Biolight	USD	FOB	121.3			
3	Cylinder - Oxygen, with regulator &flowmeter, 20L	Each	4000	1st winner:Quayle Dental & 2nd winner Fenway trading	USD	FOB	89.05	3069.8296	—	LC	as per the bid	
4	Cylinder - Oxygen, 50 L With Trolley, Regulator, Humidifier and Flow Meter	Each	550	Quayle Dental	USD	FOB	183.55	6327.53751	—	LC	as per the bid	
5	X Ray - Radiography, Digital	Each	15	Pending								
6	Ultrasound - General Purpose, Color Doppler	Each	34	Pending								
7	ECG Machine	Each	108	Guangdong Biolight	USD	FOB	1005.55	34664.43	1025.66	LC	as per the bid	
8	Bed - Patient, Adult	Each	5000	Pending								
9	Trolley - Patient, Stretcher	Each	300	Pending								
10	Defibrillator	Each	100	Pending								
1	Thermometer - Infrared, non-contact, forehead	Each	2638	1st winner: Boston medical FZE & 2nd winner Quayl Dental	USD	FOB	34.43	1186.909	—	LC	as per the bid	

የዋናው ሰነድ ከመጠቀም በፊት ትክክለኛ መሆኑን ጠቁሙ ጠቁሙ ጠቁሙ
 ke sure that this is the correct issue be



OF for Official winner list OF/TM/HO-16, 0, 13/09/2019




Company Name:		Document No: OF/TM/ HO-17
የኢትዮጵያ መድሃኒት አቅርቦት ኤጀንሲ		
Ethiopian Pharmaceuticals Supply Agency		Effective date : 18/06/2020
Reasons for Rejection		Rev. No.: 0
		Page 1 of 3

Tender no.: ICB/PSA6/COVID-19/MOH/MI/08/20

S.No	Product Description	Name of Rejected bidders	Reason for Unsuccessfulness	Remark
1	Ventilator - ICU, Mechanical	China National ins.	No Synchronized nasal CPAP	
		Esky	No valid EFDA product certificate	
		Ambulanc	No valid EFDA product certificate	
		Beevejay	The accomedation cost for installation/ training out of Addis Ababa is not quoted	
		Tecme	No valid EFDA product certificate, & Accessories not fully quoted	
		Quality Medical FZC	the product model is not as per requested, not a fully featured critical care ventilator, No valid EFDA/CE/FDA product certificate	
		Saim Assefa Ainalem importer	No Valid EFDA/CE/FDA product certificate	
		Masimo int.	No valid EFDA product certificate	
2	Pulse Oximeter	Shiller	No valid EFDA/CE/FDA product certificate	
		Quayl dental	No Valid EFDA/CE/FDA product certificate	
		Contec medical	Data storage is not indicated on catalogue	
		Ambulanc	No Valid EFDA/CE/FDA product certificate	
		Beevejay	No Valid EFDA product certificate, Battery capacity 8hr against 12hr	
		Edan	Price Incomptency	
		Shenzhen Mindray	Price Incomptency	
		Boston Medical FZE	Price Incomptency	
		Saim Assefa Ainalem importer	No Valid EFDA/CE/FDA product certificate	
		Masimo int.	No valid EFDA product certificate	



3	Cylinder - Oxygen, with regulator & flowmeter, 20L	Appotecharis	No valid CE/FDA product certificate for oxygen regulator and flowmeter, No Catalogue attached	
		Ambulanc	No valid CE/FDA product certificate for oxygen regulator and flowmeter, No model quoted, No catalogue with specification for regulator	
		Kay and Company	No valid CE/FDA product certificate for oxygen regulator and flowmeter, No Catalogue attached	
		Quality Medical FZC	No valid CE/FDA product certificate for oxygen regulator and flowmeter	
		Shanghai Eternal faith industry	No valid CE/FDA product certificate for oxygen regulator and flowmeter	
4	Cylinder - Oxygen, 50 L With Trolley, Regulator, Humidifier and Flow Meter	Appotecharis	No valid CE/FDA product certificate for oxygen regulator and flowmeter, No Catalogue attached	
		Ambulanc	No valid CE/FDA product certificate for oxygen regulator and flowmeter, No model quoted	
		Kay and Company	No valid CE/FDA product certificate for oxygen regulator and flowmeter, No Catalogue attached	
		Fenway trading	Price Incomptency	
		Quality Medical FZC	No valid CE/FDA product certificate for oxygen regulator and flowmeter	
		Shanghai Eternal faith industry	No valid CE/FDA product certificate for oxygen regulator and flowmeter	
7	ECG Machine	Shenzen Biocare biomed.	No list of spare part with their part number & costing	
		Saim Assefa Ainalem importer	No valid EFDA product certificate	
		Shiller	Price Incomptency	
		Shenzhen Mindray	Price Incomptency	
		Boston Medical FZE	Price Incomptency	
		Edan	Price Incomptency	
		Esky	No list of spare part with their part number & costing ,No gel	
		Ambulang	No Valid EFDA/CE/FDA product certificate	
		Shenzhen Comen	No Valid CE/FDA product certificate	
		Quality Medical FZC	No list of spare part, No valid EFDA product certificate	



		Contec medical	No USB, No cap availability of transfer ECG data to PC, ECG thermal paper 20m against 30m, No preview access prior to print	
11	Thermometer - Infrared, non-contact, forehead	Saim Assefa Ainalem	No Valid CE/FDA product certificate, Measurment distance : 1-5cm against 5-15 cm	
		LMB technology Gmbh	Measuring distance: 5-8cm against 5 - 15cm	
		Beevejay	Measuring distance: 10 against 5-15cm	
		Fenway trading	Measuring distance: 3-5cm against 5-15cm	
		Tianjin Orimed	Measuring distance: 5-8cm against 5-15cm, No Valid CE/FDA product certificate, No catalogue	
		Ambulanc	Measuring distance: 3-5cm against 5-15cm, No Valid CE/FDA product certificate	
N.B	Sr.No	Supplier	Remark	
	1	Zhejiang Med.	50% advance payment requested	
	2	Jupitor trading	30% advance payment requested, Local bidder quoted in foreign currency 30% advance payment requested, Local bidder quoted in foreign currency	
	3	BPL	bid bond validity 148 days against 127days	
	4	Citrus international trading	Local bidder quoted in foreign currency	
	5	Huntleigh Health care.	Delivery date more than 22weeks against 8 weeks	
	6	Jianhu Kangbo medical tech.	bid bond validity 120 days against 148 days	
	7	Hamilton	25 quantity on 1st week of October & 375 quantity on November against 60 days	

B for CE Declaration and other kind of certificates which is not stated in the bid document ITB 1.2a is not accepted.

ከዎ በዚህ ሰነድ ከመጠቀምዎ በፊት ትክክለኛ መሆኑን ያረጋግጡ
 Make sure that this is the correct issue before use

(Handwritten signatures)

