

ቀን 10 AUG 2020
Date

ቁጥር
Ref. No. EPISA/06/061/04

TO: All Participants of Tender no. ICB/PSA6/SDG& Other/Placement/LR/32/19


Dear Sirs,

Ref: - Partial Tender Result

Assessment of the above captioned tender is now partially finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceutical Supply Agency (PSA) Director General Office from August 10, 2020, to August 14, 2020; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards,


Seifu Isa Binagde
Pharmaceuticals and Medical
Supplies Tender Management
Directorate Director



ENC: 4 page

CC

- Director General
- Deputy Director General /Procurement/
- Chemical & Medical Supplies Procurement Team

PSA



Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አ.አ.አ

Document No: OETM/HO-16

Title:

Ethiopian Pharmaceuticals Supply Agency

Effective date : 13/09/2019

Official Winner list (partial)

Page 1 of 1

For tender no.: ICB/PSA6/SDG&other/Placement/LR/32/19

Item No.	Description	Unit	Qty	Winner		Current Price	Incoter m	Unit Price	Payment Terms	Delivery Time	Remark
				Supplier	Country						
Lot 1	Analyzer - Blood Grouping, Fully Automated (reagents/consumables)	Test	3,140,363	Immucor	Germany	USD	FOB	0.7188434656	LC	As per bid	
	Microplates (barcoded)	100 plates						114.62	LC	As per bid	100plates/1100 test for (ABO grouping & RH typing&Reverse grouping), 100plates/2000test for IgG phenotyping, 100plates/2300test for high titre agglutinins (Anti-A (IgM)) or (Anti-B (IgM))
	Diluent	3x500ml						117.22	LC	As per bid	1500ml/4000test for (ABO grouping&RH typing&Reverse grouping), 1500ml/4500test for high titre agglutinins (Anti-A (IgM)) or (Anti-B (IgM))
	Immucelone Anti-A IgM	10x10ml						354.28	LC	As per bid	100ml/7000test
	Immucelone Anti-B IgM	10x10ml						354.28	LC	As per bid	100ml/7000test
	Immucelone Anti-AB IgM	10x10ml						354.28	LC	As per bid	100ml/7000test
	Immucelone RH-Hr Control	10x10ml						147.62	LC	As per bid	100ml/4000test
	Immucelone Anti-D rapid IgM	10x10ml						354.28	LC	As per bid	100ml/7000test
	Novacelone Anti-D IgM+ IgG Monoclonal blend	10x10ml						354.28	LC	As per bid	100ml/7000test
	Reference cells - 2 (Group A1 and B)	2x10ml						29.52	LC	As per bid	10ml/600test for (ABO grouping&RH typing&Reverse grouping), 10ml/200test for weak D
	Capture-R* Select	480 wells						104.2	LC	As per bid	480wells/220test for weak D, 480wells/60test for IgG phenotyping
	Capture-R* Indicator Red cells	10x11.5ml						85.1	LC	As per bid	115ml/790test for weak D, 115ml/700test for antibody screening, 115ml/105test for antibody identification, 115ml/211test for IgG phenotyping
	Capture LISS	10x11.5ml						72.94	LC	As per bid	115ml/840test for weak D, 115ml/900test for antibody screening, 115ml/128test for antibody identification, 115ml/520test for IgG phenotyping
	Capture-R* Ready-Screen (I and II)	240ul						156.3	LC	As per bid	240ul/210test

[Handwritten signature]

[Handwritten signature]


Item No.	Description	Unit	Qty	Supplier	Country	Currentcy	Incoter m	Unit Price	Payment Terms	Delivery Time	Remark
	Capture-R* positive control serum (weak), Capture-R* negative control serum	2x11.5ml						79.89	LC	As per bid	11.5ml/200test for antibody screening, 11.5ml/400test for antibody identification
	Capture-R* Ready-ID	30ul						104.2	LC	As per bid	30ul/30test
	Negative control micro	5ml						21.26	LC	As per bid	5ml/140test
	Anti-Fy(a) micro	5ml						132.85	LC	As per bid	5ml/140test
	Anti-Fy(b) micro	5ml						177.14	LC	As per bid	5ml/140test
	Anti-S micro	5ml						132.85	LC	As per bid	5ml/140test
	Anti-s micro	5ml						132.85	LC	As per bid	5ml/140test
	Anti-k micro	5ml						177.14	LC	As per bid	5ml/140test
	Gammazyme-B	10ml						31.26	LC	As per bid	10ml/650test
	Monoclonal control	10x10ml						147.62	LC	As per bid	100ml/7000test
	Gammaclone Anti-Jka	5ml						260.5	LC	As per bid	5ml/400test
	Gammaclone Anti-Jkb	5ml						260.5	LC	As per bid	5ml/400test
	CoroCTM Extend Standard	10 ml						39.94	LC	As per bid	10ml/180test for weak D, used as for the system
	System liquid concentrate	10x500ml						104.2	LC	As per bid	Avg. 3-5ml/test
	Rely+onTMPer asafeTM	10 pcs						45.15	LC	As per bid	1pcs/3month/instrument
	Stir balls	2x50 pcs						50.36	LC	As per bid	1pcs used for one vial of reference cell/one vial of Capture-R* indicator Red cells
	Stir balls dispenser	pcs						199.72	LC	As per bid	2pcs/instrument/year

Lot II: Transfusion Transmissible Infections Testing Automated

Item No.	Description	Unit	Qty	Winner		Currentcy	Incoter m	Unit Price	Payment Terms	Delivery Time	Remark
				Supplier	Country						
Lot II	Transfusion Transmissible Infections Testing Automated (Reagents/Consumables)	Test	3,349,720	Abbott	Germany	USD	FOB	1,473564926	LC	As per bid	Air freight for reagents/consumables is FOC (free of charge)
	HBsAg reagent kit	kit						231.81	LC	As per bid	200 test/kit
	HBsAg Calibrator	3ml						39.90	LC	As per bid	150ul/test
	HBsAg Control	8ml						40.20	LC	As per bid	150ul/test
	HIV Ag/Ab reagent kit	kit						246.78	LC	As per bid	200 test/kit
	HIV Ag/Ab Calibrator	3ml						43.90	LC	As per bid	150ul/test
	HIV Ag/Ab Control	8ml						44.00	LC	As per bid	150ul/test
	ANNT-HCV reagent kit	kit						450.40	LC	As per bid	200 test/kit
	ANNT-HCV Calibrator	3ml						40.90	LC	As per bid	150ul/test
	ANNT-HCV Control	8ml						41.00	LC	As per bid	150ul/test
	Syphilis reagent kit	kit						216.61	LC	As per bid	200 test/kit
	Syphilis Calibrator	3ml						43.60	LC	As per bid	150ul/test
	Syphilis Control	8ml						43.70	LC	As per bid	150ul/test

Handwritten signature

Handwritten signature

	Company Name:		የኢትዮጵያ መድኃኒት አቅርቦት ኤጀንሲ		Document No: OF/TM/ HO-17
	Title:		Ethiopian Pharmaceuticals Supply Agency		Effective date : 13/09/2019
	Reasons for Rejection				Rev. No.: 0
					Page 1 of 1
Tender no. ; ICB/PSA6/SDG&other/Placement/LR/32/19					
S. No.	Item Name	Name of rejected Bidders	Item no 's rejected	Reasons for rejection	Remark
1	Analyzer - Blood Grouping, Fully	Bio-rad laboratories	Lot I	No EFDA product certificate	
2	Transfusion Transmissible Infections Testing Automated	Bio-rad laboratories	Lot II	No EFDA product certificate, double step dilution against single step dilution, no real time visual display	
		Diasorin		No EFDA product certificate, Bidder didn't do interfacing with Blood Safety Information System in the blood bank service, Bidder didn't provide introductory kits for training and equipment validation	
		SNIBE		Sample type; only serum against serum and plasma	

Others Price incompetency

