

ቀን 13 APR 2021
Date
ቁጥር
Ref. No. EPSA/06/706/04

TO: All Participants of Tender no. ICB/PSA6/RDF & SDG/MI/14/20


Dear Sirs,

Ref: - - Partial Tender Result

Assessment of the above captioned tender is now partially finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceuticals Supply Agency (PSA) Director General Office from April 13, 2021, to April 19, 2021; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards,


Tibebu Hailu Mogessie
Pharmaceuticals & Medical
Supplies Procurement Tender
Management Directorate
PIDirector




ENC: 3 page

CC

- Director General
- Deputy Director General /Procurement/
- Medical Instrument Procurement Team

PSA



Company Name:		Document No: OF/TM/ HO-16
የኢትዮጵያ መድሃኒት አቅርቦት ኤጀንሲ		
Title:		Effective date : 13/09/2019
Ethiopian Pharmaceuticals Supply Agency		
Official Winner list (Partial)		Rev. No.: 0
		Page 1 of 1

Tender no.: ICB/PSA6/RDF&SDG/MI/14/20

Item NO	Description	Unit	Qty	Winner		Currency	Unit Price	Installation and Training Cost in ETB	Incoterm	Delivery Time	Remark
				Supplier	Country						
1	X Ray - Radiography, Digital without fluoroscopy	Each	19								Pended
2	Table - Operating, Electrohydraulic	Each	134								Pended
3	Refrigerator - General Purpose, 700L	Each	40								Pended
4	Tube Sealer	Each	41								Pended
5	Bed - ICU, Electrohydraulic	Each	170								Pended
6	Couch - Blood Donor	Each	60								Pended
7	CTG Machine	Each	17								Pended
8	Ultrasound - Ob/Gyn	Each	22								Pended
9	Ultrasound - General purpose, Color Doppler, Mobile	Each	35								Pended
10	X Ray - Radiography, Digital with fluoroscopy	Each	4								Pended
11	Defibrillator - AED	Each	15								Pended
12	Laundry Machine	Each	34								Pended
13	Warmer - Radiant, Infant	Each	86								Pended
14	Refrigerator - Blood Bank, 350L	Each	143	Trifa Care Supplies Trading Inc.	Canada	USD	5,053.00	9,000.00	FOB	As per bid	
15	Refrigerator - Laboratory, 300L	Each	9								Pended
16	Freezer - Deep, 350L	Each	6								Pended
17	ECG Machine	Each	29								Pended
18	Bed - Infant/ Neonate	Each	750								Pended
19	Hammer - Reflex, 20cm	Each	120								Pended
20	Hemocytometer	Each	300								Pended
21	Humby - Knife with Blade	Each	20								Pended
22	Light - Operating, Mobile	Each	270								Pended
23	Set - Diagnostic, EENT	Set	150								Pended
24	Bed - Patient, Fowler, Adult	Each	5000								Pended
25	Trolley - Patient, Strecher	Each	300								Pended
26	Freezer - Plasma, 550L	Each	100	Trifa Care Supplies Trading Inc.	Canada	USD	8,400.00	9,000.00	FOB	As per bid	
27	Plasma Thawing Device	Each	30	Trifa Care Supplies Trading Inc.	Canada	USD	5082.00	5,000.00	FOB	As per bid	
28	Warmer - Blood	Each	225	Spot Medical Equipment FZE.	U.A.E	USD	1280.00	Included	FOB	As per bid	
29	Thermometer - (Min-Max)	Each	500								Cancelled
30	Water Bath	Each	70								Cancelled
31	Cold Box	Each	250								Cancelled
32	Centrifuge - Refrigerated	Each	10								Cancelled
34	Set - Chest tube	Set	60								Pended
35	CPAP	Each	60								Pended
36	Set - Intraosseous	Set	60								Pended
37	Trolley - Mayo type	Each	164								Pended
38	Set - Intubation	Set	20								Pended

አባኩዎ በዚህ ሰነድ ከመጠቀምዎ በፊት ትክክለኛ መሆኑን ያረጋግጡ
 Make sure that this is the correct issue before use

OF for Official winner list OF/TM/HO-16, 0, 13/09/2019



Company Name:		Document No: OF/TM/ HO-17
	የኢትዮጵያ መድሃኒት አቅርቦት ኮሚሽን	
Title:		
	Ethiopian Pharmaceuticals Supply Agency	Effective date: 13/09/2019
	Reasons for Rejection (Partial)	Rev. No.: 0
		Page 1 of 1


Tender no.: ICB/PSA6/RDF&SDG/MI/14/20

Preliminary Unsuccessful Bidders

S.No.	Name of Bidder	Reason of unsuccessfulness
1	C.F.DI Ciro Fiocchetti & C	Delivery time; 4-8 months against 60-90 days

Technically Unsuccessful Bidders

Item no	Description	Unsuccessful Bidder	Reason of unsuccessfulness
14	Refrigerator - Blood Bank, 350L	Qingdao Haier Biomedical Co., Ltd	No CE/US FDA product certificate from SRA regions
		Rise Import and Export GmbH	No CE/US FDA product certificate from SRA regions, not touch screen, Noise level not indicated, No list of important spare part with part no & costing, No five year maintenance contract with price break down
		CTEQ Limited	Capacity: 96 bags of 450ml against 170 bags of 450ml, or 350L against 303L, Noise level not indicated, No five year maintenance contract with price break down, Model mismatch between catalogue and quotation.
		Apothecares	No CE/ US FDA product certificate from SRA regions, Auto-defrosting not indicated, Integrated chart recorder not indicated, CFC free not indicated, Blood temperature control range; +4 to +6 °C against +2 to +6 °C
		ESCO	No CE/ US FDA product certificate from SRA regions, No list of important spare part with part no & costing, No five year maintenance contract with price break
		International Consolidated Contractors	No CE/ US FDA product certificate from SRA regions, Not touch screen
		Boston Medical FZE	No CE/US FDA product certificates from SRA region, Single glass door against dual glass door
		B Medical Systems	No SRA product certificate, Manual closing door against Self closing door if accidentally left opened less than 20°
		Unique Enterprise Limited	No CE/US FDA product certificate from SRA regions, Scree size; 4.5" against 6.5"
		EN-Project	No CE/US FDA product certificate from SRA regions, Capacity; 280L against 350L or 165 bag of 450ml against 170 bag of 450ml
26	Freezer - Plasma, 550L	Qingdao Haier Biomedical Co., Ltd	No EFDA product certificate, No CE/US FDA product certificate from SRA regions, No touch screen, Rack material; stainless steel against transparent poly carbonate
		Rise Import and Export GmbH	NO CE/US FDA product certificate from SRA regions, manual deforsting against auto-defrosting, Rack material not indicated, No list of important spare part with part no. & costing, No five year maintenance contract with price break dow, Noise level not indicated
		CTEQ Limited	No EFDA product certificate, No voltage stabilizer, Rack material: stainless steel against transparent poly carbonate, Temperature accuracy: +/-5 dgeree centigrade against +/-1 dgeree centigrade
		ESCO	No EFDA product certificate, No CE/US FDA product certificate from SRA regions, Noise level; 60db against <50db, Rack material type not indicated, temperature control range -20 to -40 degC against -18 to -30degC, No two year
		International Consolidated Contractors	No CE/US FDA product certificate from SRA regions, Battery capacity not indicated
		Boston Medical FZE	No EFDA product certificate, No CE/US FDA product certificate from SRA regions, Rack materila; stainless steel against transparent polycarbonate, Battery capacity; 8hr against 18hrs
		B Medical Systems	No EFDA product certificate, No CE/US FDA product certificate from SRA region, Manual closing door against Self closing door incase of accidentally opened door

		Unique Enterprise Limited	No CE/US FDA product certificate from SRA regions, Manual defrosting against auto-defrosting, Rack material; stainless steel against transparent polycarbonate
		EN Project	No EFDA product certificate, No CE/US FDA product certificate from SRA regions, Rack material; Aluminium against transparent poly carbonate, Internal material type; Aluminium against stainless steel
27	Plasma Thawing Device	International Consolidated Contractors	No EFDA product certificate, No CE/US FDA product certificate from SRA regions
		Unique Enterprise Limited	No CE/US FDA product certificate from SRA regions, Plasma thawing capacity; 6 bags at a time against 8 bags at a time, No list of important spare parts with part no and costing
28	Warmer - Blood	Keewell Medical Technoloy Co.,Ltd	No CE/US FDA product certificate from SRA regions
		CTEQ Limited	NO list of important spare part with part no. and costing, No five years maintenance contract with price break down, No spare battery pack
		Shenzhen Bestman	No CE/US FDA product certificate from SRA regions, No two years basic warranty, No five years maintenance contract with price break down, No list of important spare parts with part no. and costing, No Spare battery pack
29	Thermometer - (Min-Max)	CTEQ Limited	No CE/ US FDA product certificate from SRA regions
30	Water Bath	Apothecares	Accessories not quoted, No CE/US FDA product certificate from SRA regions
		Boston Medical FZE	No CE/US FDA product certificate from SRA regions, No list of important spare parts
		Unique Enterprise Limited	No CE/US FDA product certificate from SRA regions
		CTEQ limited	No CE/US FDA product certificate from SRA regions, No wassermann tubes
		Spot Medical FZE	No CE/US FDA product certificate from SRA region
31	Cold Box	CTEQ Limited	No WHO product certificate or No ISO 13485 product certificate from SRA region, Cold life; 10/17hrs against 130hrs in ambient temperature +43 degC
		Apothecares	No WHO product certificate or No ISO 13485 product certificate from SRA region
		B Medical Systems	Cold life; 56.5hrs against maintenance of under +10degC for 130 hours in an ambient temperature of +43degC
		Unique Enterprise Limited	No WHO product certificate or No ISO 13485 product certificate from SRA region
32	Centrifuge - Refrigerated	CTEQ Limited	No CE/US FDA product certificate from SRA regions, RCF; 3350g against 7200g, Speed; 4500rpm against 5000rpm, Timer; 10s - 999min against 0 - 99h 59min
		Apothecaries	No CE/US FDA product certificate from SRA regions, Model not indicated, Swing-out rotor; 12x15ml against 12x800ml
		Boston Medical FZE	No CE/US FDA product certificate from SRA regions, Speed; 4700rpm against 5000rpm
		Unique Enterprise Limited	No CE/US FDA product certificate from SRA regions,Timer; 10s - 999min against 0 - 99h 59min, No CE/US FDA product certificate from SRA regions, Speed; 4400rpm against 5000rpm
			
አባዘዎ በዚህ ሰነድ ከመጠቀምዎ በፊት ትክክለኛ መሆኑን ያረጋግጡ። Make sure that this is the correct issue before use		OF for Reasons for Rejection, OF/TM/HO-17, 0, 13/09/2019	