

ቀን 25 JUN 2021
Date
ቁጥር
Ref. No. EPISA/66/912/04

TO: All Participants of Tender no. ICB/PSA6/RDF & SDG/MI/14/20

Dear Sirs,

Ref: - 2nd Partial Tender Result

Assessment of the above captioned tender is now partially finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceuticals Supply Agency (PSA) Director General Office from June 26, 2021, to July 1, 2021; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards,


Hailu Mogessie
Pharmaceuticals & Medical
Supplies Procurement Tender
Management Directorate
P/Director



ENC: 13 page

CC

- Director General
- Deputy Director General /Procurement/
- Medical Instrument Procurement Team


PSA

Company Name: የኢትዮጵያ መድኃኒት አቅርቦት ሎጂስቲክስ
 Title: Ethiopian Pharmaceuticals Supply Agency
 Document No: OF/TM/HO-15
 Effective date : 13/09/2019

Winner list (Partial)
 Tender no.; ICB/PSA6/RDF&SDG/ML/14/20

Item NO	Description	Unit	Qty	Winner		Currency	Unit FOB Price		Installation and Training Cost in ETB	Turnkey cost in ETB	Incoferm	Delivery Time	Remark
				Supplier	Country		Foreign	Birr					
1	X Ray - Radiography, Digital without fluoroscopy	Each	19	Ertunc Ocean	Turkey	USD	130765	5659979.35	Included	Included	FOB	As per bid	
2	Table - Operating, Electrohydraulic	Each	161	Retendered									
3	Refrigerator - General Purpose, 700L	Each	40	Trifa Care Supplies Trading Inc.	Canada	USD	6136	248485.91	9000	NA	FOB	As per bid	
4	Tube Sealer	Each	41	Cancelled									
5	Bed - ICU, Electrohydraulic	Each	170	Cancelled									
6	Couch - Blood Donor	Each	60	L_inet		EUR	2174	106782.75	Included	NA	FOB	As per bid	
7	CTG Machine	Each	17	Cancelled									
8	Ultrasound - Ob/Gyn	Each	22	Pended									
9	Ultrasound - General purpose, Color Doppler, Mobile	Each	35	Cancelled									
10	X Ray - Radiography, Digital with fluoroscopy	Each	4	Pended									
11	Defibrillator - AED	Each	15	Schiller Ag		USD	5790	234474.16	Included	NA	FOB	As per bid	
12	Laundry Machine	Each	34	Cancelled									
13	Warmer - Radiant, Infant	Each	86	Cancelled									
15	Refrigerator - Laboratory, 300L	Each	9	Pended									
16	Freezer - Deep, 350L	Each	6	Cancelled									
17	ECG Machine	Each	29	Cancelled									
18	Bed - Infant/ Neonate	Each	750	Cancelled									
19	Hammer - Reflex, 20cm	Each	120	Cancelled									
20	Hemocytometer	Each	300	Cancelled									
21	Humby - Knife with Blade	Each	20	Cancelled									
22	Light - Operating, Mobile	Each	324	Boston Medical FZE		USD	7911.7	320395.37	10000	NA	FOB	As per bid	
23	Set - Diagnostic, EENT	Set	150	Cancelled									
24	Bed - Patient, Fowler, Adult	Each	5000	Cancelled									
25	Trolley - Patient, Stretcher	Each	300	Cancelled									
34	Set - Chest tube	Set	60	Cancelled									
35	CPAP	Each	60	Cancelled									
36	Set - Intraosseous	Set	60	Cancelled									
37	Trolley - Mayo type	Each	164	Cancelled									
38	Set - Intubation	Set	20	Cancelled									

Others; Price incompetency



OF for Winner list OF/TM/HO-15, 0, 13/09/2019

እርስዎ በዚህ ስነ፡ ኮምፒዩተር በፊት ትክክለኛ መሆኑን ያረጋግጡ
 Make sure that this is the correct issue before use


Tender no.: ICB/PSA6/RDF&SDG/MI/14/20

Preliminary Unsuccessful Bidders (Partial)

S.No.	Name of Bidder	Reason of Rejection
1	C.F.DI Ciro Fiocchetti & C	Delivery time; 4-8 months against 60-90 days
2	Supra Health Care Johannesburg Pvt. Ltd.	Delivery time: for item no. 5, 18, 24 and 25; 120 days against 60-90 days
3	Ethio-Mecan Import Trade.	Payment term; 40% in advance and 60% after delivery
4	Shanghai United Imaging Health care Co.,Ltd	Payment Term; 100% upon shipment against 90% upon shipment &10% upon commissioning of devices
5	Heal Force international Trading Co.,Ltd.	Bid bond validity; 120 days against 148 days

Technically Unsuccessful Bidders (Partial)


Item no	Description	Unsuccessful Bidders	Reason of unsuccessfulness
1	X Ray - Radiography, Digital without fluoroscopy	Beijing Wandong Medical Equipment	- No vertical bucky having tilting facility across +90 degrees
		Shimadzu Corporation Cteq Ltd	- No facility of tracking with x ray tube - KV; 40-125kv against 40-150KV, - mA; 10-600mA against 10-800mA, Power; 50KW against 65-80KW, No CE/US FDA product certificate from SRA region
		Rise Import and Export Gmbh	- No manufacturer catalog , - Longitudinal Movement; 130cm against 210cm - No lateral movement
		Supra HealthCare Johannesburg Pvt. Ltd.	No manufacturer catalog with detail specification - No two year warranty, - Online training against onsite training
		Shenzhen Mindray Bio-medical Electronics Co.Ltd	- No UPS with 30 min back up, - No exposure switch,
		Esky Medical Company	- Warranty; 2 year from date shipment against two year from date of commissioning, - Lateral movement; 22cm against 24cm
		Elsmed Healthcare Solutions	- Active matrix size: 2.3k x 2.8k against approx. 2.8k x 2.8k, - No CE/US FDA product certificate from SRA region
		Boston Medical FZE	mA range; 10-630m against 10-800mA -Power; 50KW against 65-80KW, Anode heat capacity; 150 KHU against 300KHU - No vertical bucky having tilting facility across +90 degrees
		Trading Medical Systems Private Co.Ltd	- DQE; 58% against 65%, - Pixel size; 150um against 148um, - Patient Carrying capacity; 170kg against 200kg, - Detector size; 35x43 cm against 43x43 cm, - No CE/US FDA product certificate from SRA region
		2	Table - Operating, Electrohydraulic
Medifa Gmbh and Co. KG	No CE/US FDA product certificate from SRA region		



Heyer Medical Co., Ltd	Height adjustment; 690-950mm against 680-1100mm, No CE/US FDA product certificate from SRA region -Table; electric operated against electrohydraulic operated
Nanjing Mindray	Carrying capacity; 230kg against 250kg, No CE/US FDA product certificate from SRA region
SCHMITZ	No CE/US FDA product certificate from SRA region
Jiangsu Saikang Medical Co.,Ltd	Height adjustment; 500-800mm against 680-1100mm, No CE/US FDA product certificate from SRA region -Model miss match between catalog and quotation,
Guangdong Kangshen Medical Technology	No original manufacturer catalog No CE/US FDA product certificate from SRA region
Esky Medical Company	Carrying capacity; 200kg against 250kg No CE/US FDA product certificate from SRA region
Apothecaries Sundries	-Height adjustment; 750-1000mm against 680-1100mm, No CE/US FDA product certificate from SRA region -Accessories not fully quoted
Boston Medical FZE	-Carrying capacity; 180kg against 250kg, No CE/US FDA product certificate from SRA region
Cteq Ltd	Carrying capacity; 225kg against 250kg, No CE/US FDA product certificate from SRA region
Quayle Dental Manufacturing (Export) Ltd	-Height adjustment; 700-1000mm against 680-1100mm, No CE/US FDA product certificate from SRA region
Rise Import and Export GmbH	No manufacturer catalog for quoted model, No CE/US FDA product certificate from SRA region
Narula Export	-Height adjustment; 725-980mm against 680-1100mm, No CE/US FDA product certificate from SRA region
3 Refrigerator - General Purpose, 700L	Qingdao Haier Biomedical Co., Ltd
	-No EFDA and CE/US FDA product certificate from SRA region, -spark free ignition against without spark free ignition
	Cteq Ltd Ltd
	-No EFDA and CE/US FDA product certificate from SRA region
	Esco World Class
	-No EFDA and CE/US FDA product certificate from SRA region, -Volume; 620L against 700L, -Accuracy; 3 degc. against +/- 1 degc.
	International consolidated
	-No CE/US FDA product certificate from SRA region,



			-No recording Fetal ECG, -Screen size; 12.1" against 15"
		Boston Medical FZE	-Built-in rechargeable battery not quoted, -Accessories not fully quoted, -No storage memory capability
		Quality Medical FZC	Screen size; 6" against 15"
9	Ultrasound - General purpose, Color Doppler, Mobile	Samsung Medison Co.,Ltd	No 4D capability for linear and phased array probe
		Shenzhen Mindray Bio-medical Electronics Co.Ltd	No 4D capability for linear and phased array probe
		Quality Medical FZC	-Dynamic range; 200dB against 250dB or higher, -No 4D capability for linear and phased array probe, -No CE/US FDA product certificate from SRA region
		Rise Import and Export Gmbh	-No upgradeable to fusion/navigation, -Black & white printer against color printer, - No 4D capability for linear and phased array probe
		Spot Medical Equipment FZE	Linear probe FOV; 4-36 deg. against 40 deg.
		Elsmed Healthcare Solutions	-No 4D capability for linear and phased array probe, -No CE/US FDA product certificate from SRA region
		Boston Medical FZE	No CE/USFDA product certificate from SRA region
		Trading Medical Systems Private Co.Ltd	-Micro curved probe 2-5MHz against 4-9MHz, - FOV: 80 against 134 Phased array : 120 deg against 128deg
		Sonoscape Medical Corp.	No 4D capability
		CHISON	-No 4D capability -Phased array FOV: 84 against 128
11	Defibrillator - AED	APOTHECARIES SUNDRIES	-Thermal paper not quoted, -Manual mode charging time at max. energy; <6s against 5s, -No CE/US FDA product certificate from SRA region
		Boston Medical FZE	-Manual lead switching against automatic switching, -Manual mode charging time at max. energy; <7s against 5s, -3 and 5 lead against 12 lead, -No CE/US FDA product certificate from SRA region
12	Laundry Machine	Trifa Care Supplies Trading Inc.	No Product certificate from EFDA & SRA Region
		UNIQUE	Device uses electrical against both electrical and steam
		Qingdao Haier Biomedical Co., Ltd	-NO CE/US FDA product certificate from SRA region, -Drum volume; 515L against 530L
		TOLON GLOBAL MAKINA	NO CE/US FDA product certificate from SRA region
		International consolidated	-No manufacturer catalog,



			-NO CE/US FDA product certificate from SRA region
		Spot Medical Equipment FZE	NO CE/US FDA product certificate from SRA region
		Tolkar Smartex	NO CE/US FDA product certificate from SRA region
		Quality Medical FZC	NO CE/US FDA product certificate from SRA region
13	Warmer - Radiant, Infant	EN-Project	-No Inbuilt Compressor,
			-No countdown timer,
			-No CE/US FDA product certificate from SRA region
		RISE	-No Inbuilt Compressor,
			-No CE/US FDA product certificate from SRA region
		Narula Export	-No Inbuilt Compressor,
			-No weighing scale,
			-No alarm for all requested parameter (such as SPO2),
			-No CE/US FDA product certificate from SRA region
		APOTHECARIES SUNDRIES	-No inbuilt compressor,
			-No alarm for all requested parameter (such as SPO2),
			-No weighing scale,
		Ningbo David Medical Device Co., Ltd	-No CE/US FDA product certificate from SRA region
No inbuilt compressor			
Boston Medical FZE	No weighing scale, No inbuilt SPO2 module, NO CE/US FDA product certificate from SRA region		
Bistos Co., Ltd	-No Oxygen analyzer module,		
	-No T-piece integrated resuscitator,		
Fanem Ltda	-No CE/US FDA product certificate from SRA region		
	-No Inbuilt Compressor,		
	-No countdown timer,		
Ertunc Ozcan	-No CE/US FDA product certificate from SRA region		
	-No CE/US FDA product certificate from SRA region		
16	Freezer - Deep, 350L	The Medical Export	No Inbuilt Compressor
		Qingdao Haier Biomedical Co., Ltd	No CE/US FDA product certificate from SRA region
		Cteq Ltd	-No CE/US FDA product certificate from SRA region,
			-Installation, Technical and End User Training and Warranty not provided
		EN-Project	No CE/US FDA product certificate from SRA region
		Esco World Class	-No CE/US FDA product certificate from SRA region,
			-Accuracy : +/-3°C against +/-1°C
International consolidated	No CE/US FDA product certificate from SRA region		
Rise import und export GmbH	No CE/US FDA product certificate from SRA region		



		APOTHECARIES SUNDRIES	No CE/US FDA product certificate from SRA region
		B Medical Systems S.A.R.L	No CE/US FDA product certificate from SRA region
		Unique Enterprise Limited	No CE/US FDA product certificate from SRA region
17	ECG Machine	Shenzhen carewell electronics co., ltd	Sampling rate: 1KHz against 10KHz
		CONTEC MEDICAL SYSTEM	-Sampling rate: 1KHz against 10KHz, -No CE/US FDA product certificate from SRA region
		Narula Export	No CE/US FDA product certificate from SRA region
		APOTHECARIES SUNDRIES	No CE/US FDA product certificate from SRA region
		Boston Medical FZE	No CE/US FDA product certificate from SRA region
		Schiller AG	-Sampling rate: 1KHz against 10KHz, -Resolution; 320x240 dots against 800 x 480 dots
18	Bed - Infant/ Neonate	Mission Pharma	-Frame material type; plastic against thermal spray -No CE/US FDA or ISO 13485 product certificate from SRA region
		Tianjin Orimed International	Frame material type not indicated
		APOTHECARIES SUNDRIES	-Frame material type; epoxy powder coated against thermal spray, -No CE/US FDA or ISO 13485 product certificate from SRA region
		Linnet	Frame material type; powder coated against thermal spray
		Jiangsu Saikang Medical Co.,Ltd	Frame material type; plastic against thermal spray
		Hebei Pukang Medical Instruments Co., Ltd	Frame material type; epoxy coated against thermal spray
		Rise import and export Gmbh	-No manufacturer catalog for quoted model, -No CE/US FDA or ISO 13485 product certificate from SRA region
		Spot Medical Equipment FZE	Frame material type; epoxy coated against thermal spray
		Guandong kangshen Medical Technology	-Frame material type; epoxy coated against thermal spray, -No CE/US FDA or ISO 13485 product certificate from SRA region
		Narula Export	Catalog not original
		Boston Medical FZE	-Frame material type; epoxy powder coated against thermal spray, -No CE/US FDA or ISO 13485 product certificate from SRA region
		Ningbo David Medical Device Co., Ltd	Frame material type; epoxy coated against thermal spray
		United Poly Engineering pvt. ltd	Frame material type; epoxy coated against thermal spray
		Yamrot Engineering	-Frame material type; epoxy powder coated against thermal spray, -No braking of castors for quoted model,




			-No CE/US FDA or ISO 13485 product certificate from SRA region
		Ertunc Ozcan	No CE/US FDA or ISO 13485 product certificate from SRA region
		The Medical Export	-Copy paste of EPSA spec. and catalogue not original, -No CE/US FDA or ISO 13485 product certificate from SRA region
19	Hammer - Reflex, 20cm	Supra HealthCare Johannesburg Pvt. Ltd.	No CE/US FDA product certificate from SRA region
		Cteq Ltd	No CE/US FDA product certificate from SRA region
		Quayle Dental Manufacturing (Export)	No CE/US FDA product certificate from SRA region
20	Hemocytometer	Cteq Ltd	No CE/US FDA product certificate from SRA region
21	Humby - Knife with Blade	Cteq Ltd	No CE/US FDA product certificate from SRA region
22	Light - Operating, Mobile	Heyer Medical Co., Ltd	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region, -Diameter of light head 50cm against 60cm
		Esky Medical Company	No EFDA product certificate, -No CE/US FDA product certificate from SRA region
		EN-Project	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region,
		Nanjing Mindray	No CE/US FDA product certificate from SRA region
		Shanghai Fuhong Science Technology Dev't Co.Ltd	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region
		Dr. Mach	-No CE/US FDA product certificate from SRA region
		Spot Medical Equipment FZE	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region
		Cteq Ltd	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region,
		Apothecaries Sundries	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region,
		Narula Export	-No EFDA product certificate, -No CE/US FDA product certificate from SRA region,
		Quayle Dental Manufacturing (Export) Ltd	No EFDA and CE/US FDA product certificate
		23	Set - Diagnostic, EENT
Tianjin Orimed	-Convenient and stable handle does not include charger, -No CE/US FDA product certificate from SRA region		



		Cteq Ltd	-Specula size; 2.5mm and 9mm not quoted, -No CE/US FDA product certificate from SRA region
24	Bed - Patient, Fowler, Adult	Jiangsu Saikang Medical Co.,Ltd	-Material type; epoxy coated against thermal spray, -Side rail not adjustable, -Model miss match between catalog and quotation, -No side support to immobilize mattress
		Tianjin Orimed International	-Carrying capacity not indicated, -Material type not indicated, -No protective bumper, -No side support to immobilize mattress
		Mission Pharma	-Carrying capacity not indicated, -No CE/US FDA or ISO 13485 product certificate from SRA region, -No side support to immobilize mattress
		Linet	-Material type not indicated, -side rail not adjustable
		Hebei Pukang Medical Instruments Co., Ltd	-Material type; epoxy coated against thermal spray, -No CE/US FDA or ISO 13485 product certificate from SRA region, -No side support to immobilize mattress
		Spot Medical Equipment FZE	No manufacturer catalog for quoted model
		Rise import and export GmbH	-Carrying capacity not indicated, -No protective bumper, -No CE/US FDA or ISO 13485 product certificate from SRA region
		Guandong Kangshen Medical Technology	-No manufacturer catalog with detail specification, -Quoted model not indicated
		Narula Export	No manufacturer catalog for quoted model
		APOTHECARIES SUNDRIES	-Carrying capacity not indicated, -No protective bumper, -No CE/US FDA or ISO 13485 product certificate from SRA region, -No side support to immobilize mattress, -side rail not adjustable
		Boston Medical FZE	No CE/US FDA or ISO 13485 product certificate from SRA region
		Esky Medical Company	-Material type; epoxy coated against thermal spray, -No CE/US FDA or ISO 13485 product certificate from SRA region
		United Poly Engineering pvt. ltd	-No protective bumper, -Carrying capacity not indicated, -No side support to immobilize mattress, Side rail not adjustable
		Yamrot Engineering	-Quoted product; IV stand against patient bed, -No CE/US FDA or ISO 13485 product certificate from SRA region
Ertunc Ozcan	-Copy paste of EPSA specification, -Catalogue not original, -No CE/US FDA or ISO 13485 product certificate from SRA region		



		Trifa Care Supplies Trading Inc.	-Material type not indicated, -side rail not adjustable, -No side support to immobilize mattress
25	Trolley - Patient, Stretcher	TIANJIN INTERNATIONAL ORIMED	-Dimension; 190 x 67 x 57 cm against 205 x 75.5 cm, -No protective bumper, -No integrate holder for oxygen cylinder
		SCHMITZ	No CE/US FDA or ISO 13485 product certificate from SRA region
		Mission Pharma	-No CE/US FDA or ISO 13485 product certificate from SRA region, -Dimension; 190 x 63 x 80 cm against 205 x 75.5 cm, -Carrying capacity not indicated
		Linet	-Side rail length; 125cm against 140cm, - Mattress doesn't contain special slide-out mats for easier patient transfer
		Hebei Pukang Medical Instruments Co., Ltd	-No CE/US FDA or ISO 13485 product certificate from SRA region, -No integrated holder for oxygen cylinder
		Rise import and export Gmbh	-No CE/US FDA or ISO 13485 product certificate from SRA region, -Dimension 190 x 60 x 80 cm against 205 x 75.5 cm
		Spot Medical Equipment FZE	-No CE/US FDA or ISO 13485 product certificate from SRA region, -Dimension 190 x 65cm against 205 x 75.5 cm
		Guandong Kangshen Medical Technology	-Dimension; 1950X640X610-910 mm against 205 x 75.5cm, -No manufacturer catalog with detail specification, -Quoted model not indicated
		Narula Export	No manufacturer catalog with detail
		APOTHECARIES SUNDRIES	-No CE/US FDA or ISO 13485 product certificate from SRA region, -No protective bumper, -2 holder against 4 holder for infusion drainage bags, -Carrying capacity not indicated
		Esky Medical Company	-No CE/US FDA or ISO 13485 product certificate from SRA region, -No protective bumper, -Carrying capacity; 133Kg against 200Kg
		Boston Medical FZE	-No CE/US FDA or ISO 13485 product certificate from SRA region, -Dimension; 70 x 200 cm against 75.5 x 205
		United Poly Engineering pvt. ltd	-Carrying capacity not indicated, -Castor diameter; 150mm against 200mm
		Yamrot Engineering	-Quoted model is mayo table, -No CE/US FDA or ISO 13485 product certificate from SRA region
		Ertunc Ozcan	-No CE/US FDA or ISO 13485 product certificate from SRA region, -Copy paste of EPSA Specification
37	Trolley - Mayo type	Cteq Ltd	-No CE/ US FDA product certificate from SRA region,



		-Ground clearance 75cm against 100cm
SCHMITZ		No CE/ US FDA product certificate from SRA region
Tianjin Orimed International		-No CE/ US FDA product certificate from SRA region, -Height adjustment and Frame diameter not indicated
APOTHECARIES SUNDRIES		-No CE/ US FDA product certificate from SRA region, -Tray dimension; 650 x 400 mm against 550 x 450 mm
Jiangsu Saikang Medical Co.,Ltd		-No CE/ US FDA product certificate from SRA region, -Tray dimension; 500 x 350 mm against 550 x 450 mm
Hebei Pukang Medical Instruments Co., Ltd		-No CE/ US FDA product certificate from SRA region, -Tray dimension height; 20 mm against 200
Rise Import and Export Gmbh		-No CE/ US FDA product certificate from SRA region -Copy paste of EPSA specification
Guandong Kangshen Medical Technology		-No CE/ US FDA product certificate from SRA region, -Tray dimension; 450 x 350 mm against 550 x 450 mm
Narula Export		-No CE/ US FDA product certificate from SRA region, -Castor diameter; 0.5 cm against 10 cm
Boston Medical FZE		-No CE/ US FDA product certificate from SRA region, -Tray dimension; 520 x 350 mm against 550 x 450 mm
United Poly Engineering pvt. ltd		-No CE/US FDA product certificate from SRA region, -Tray dimension; 540 x 420 mm against 550 x 450 mm, -Material type: Epoxy powder against stainless steel, -Castors has no brake
Yamrot Engineering		Material type : Epoxy powder against stainless steel
Ertunc Ozcan		-No CE/ US FDA product certificate from SRA region, -Copy paste of EPSA specification



[Handwritten signature]

			-Volume; 638L against 700L
		Rise import and export Gmbh	-No CE/US FDA product certificate from SRA region, -Key operated door lock against pin security door lock
		Boston Medical FZE	-No EFDA and CE/US FDA product certificate from SRA region, -Key operated door lock against pin security door lock
		B Medical Systems S.A.R.L	-No EFDA and CE/US FDA product certificate from SRA region, -Volume; 636L against 700L
		Supra HealthCare Johannesburg Pvt. Ltd.	-No EFDA and CE/US FDA product certificate from SRA region, -Keeping inside temperature during power failure; 2hr against 8 hr
		Lmb Technologie GMBH	-No EFDA product certificate, -Volume; 670L against 700L, Accuracy; +/-2 deg. C against +/-1 deg.C, Door type; Glass against solid
		Unique Enterprise Limited	No EFDA product certificate and CE/US FDA product certificate from SRA region
4	Tube Sealer	Boston Medical FZE	No CE/US FDA product certificate from SRA region
		Lmb Technologie GMBH	-No two year warranty, -User training not included, -Simultaneous number of seal not indicated, - No. of seal per charge not indicated
		Unique Enterprise Limited	-Device not battery operated, -No CE/US FDA product certificate from SRA region
5	Bed - ICU, Electrohydraulic	Tianjin Orimed International	No original manufacturer catalogue
		Jiangsu Saikang Medical Co.,Ltd	-No manufacturer catalog for bedside cabinet and over bed table, -Trendelenburg /reverse trendelenburg; 12 deg. against 15 deg.
		Linnet	-Electrical operated against electro hydraulic operated, -No x ray cassette holder
		Hebei Pukang Medical Instruments Co., Ltd	-Electrical operated against electro hydraulic operated, -NO CE/US FDA or ISO 13485 product certificate from SRA region, -material type; epoxy coated against electronic thermal spray
		Rise Import and Export Gmbh	-Trendelenburg/reverse trendelenburg; 12 deg. against 15 deg., -Electrical operated against electro hydraulic operated, -NO CE/US FDA or ISO 13485 product certificate from SRA region
		Narula Export	-Electrical operated against electro hydraulic operated, -Trendelenburg/reverse trendelenburg; 12 deg. against 15 deg.,



(Handwritten signature)

			-material type; epoxy coated against electronic thermal spray
		Spot Medical Equipment FZE	-NO CE/US FDA or ISO 13485 product certificate from SRA region, -Material type; epoxy painted against electronic thermal spray
		Guandong Kangsheng Medical Technology	-No catalog with detail specification, -NO CE/US FDA or ISO 13485 product certificate from SRA region, -Quoted model not indicated
		APOTHECARIES SUNDRIES	-Electrical operated against electro hydraulic operated, -NO CE/US FDA or ISO 13485 product certificate from SRA region
		Boston Medical FZE	-Electrical operated against electro hydraulic operated, -NO CE/US FDA or ISO 13485 product certificate from SRA region, -No facility to hold patient monitors.
		Esky Medical Company	-Bed Sour Mattress: Including Electrically Operated Pressure Mattress not quoted, - NO CE/US FDA or ISO 13485 product certificate from SRA region
		Ertunc Ozcan	-NO CE/US FDA or ISO 13485 product certificate from SRA region, -Copy paste of EPSA specification, -Catalog not original
		Trifa Care Supplies Trading Inc.	Electrical operated against electro hydraulic operated
6	Couch - Blood Donor		
		Rise Import and Export Gmbh	-NO CE/US FDA or ISO 13485 product certificate from SRA region, -Arm rests have swinging out as well as up and down moving facility not indicated
		APOTHECARIES SUNDRIES	-In case of vasovagal attack, donor's head neither lowered immediately nor legs lifted above his heart level, -NO CE/US FDA or ISO 13485 product certificate from SRA region
		Boston Medical FZE	No Remote control
		Esky Medical Company	-No reclining and upright body positions with a smooth shifting to any position, -No up/down control, -No electrically operated tilting function to head low position, -NO CE/US FDA or ISO 13485 product certificate from SRA region
		Ertunc Ozcan	-Copy paste EPSA specification, -Catalog not original, -NO CE/US FDA or ISO 13485 product certificate from SRA region
		Lmb Technologie GMBH	-No manufacturer catalogue, -NO CE/US FDA or ISO 13485 product certificate from SRA region
7	CTG Machine	APOTHECARIES SUNDRIES	-Ultrasound working frequency; 1MHz against 2MHz,

