

ቀን
Date 10/2 NOV 2021.
ቁጥር
Ref. No. EP/SA/06/407/04

TO: All Participants of Tender no. ICB/EP/SA6/SDG/MI/12/21

Dear Sirs,

Ref: - Tender Result

Assessment of the above captioned tender is now finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceutical Supply Agency (EP/SA) Director General Office from November 12, 2021, to November 18, 2021; 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards,


Tibebe Hailu Mogessie
Ethiopian Pharmaceuticals & Supplies Procurement Management Agency
Director General


ENC: 2 page

CC

- Director General
- Deputy Director General /Inbound Logistic/
- Medical Instrument Procurement Team

EP/SA



Company Name: የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት
 Title: Ethiopian Pharmaceuticals Supply Agency
 Document No: OF/TM/ HO-16
 Effective date : 13/09/2019

Official Winner list
 Rev. No.: 0
 Page 1 of 1

Tender no.: ICB/EPSA6/SDG/MM/12/21

Item No	Description	Unit	Qty	Winner		Currency	Unit Price	Installation and Training Cost in ETB	Turnkey cost in ETB	Incoterm	Delivery Time	Remark
				Supplier	Country							
1	Table – Operating, Electro hydraulic	Each	170	Jiangsu Saikang Medical Co., Ltd.	China	USD	14,500.00	5,120.00	NA	FOB	60 days	

N.B:- The others are Price Incompetency

(Handwritten signatures)



ለቅድመ ገቢ ላይ ከመጠየቅ በፊት ትክክለኛ መሆኑን ያረጋግጡ
 Make sure that this is the correct issue before use

OF for Official Winner list OF/TM/HO-16, 0, 13/09/2019



Tender no: ICB/EP/SA6/SDG/MI/12/1

Preliminary unsuccessful bidders

Company Name:		የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት	Document No: OF/TM/ HO-17
Title:		Ethiopian Pharmaceuticals Supply Agency	Effective date : 13/09/2019
Reasons for Rejection			Rev. No.: 0
			Page 1 of 1

S.No.	Name of Bidder	Reason of Unsuccessfulness
1	Heyer Medical Co.,Ltd.	Bid Price Validity: 90 days against 120 days, Payment term: 100% against 90/10%
2	Apothecares Sundries Mfg. Co.	Bid bond validity: 121 days against 148 days
3	Boston Medical FZE	Bid bond validity: 121 days against 148 days, Delivery time: 120-150 days against 60-90 days

Technically unsuccessful bidders

Item no	Description	Unsuccessful Bidder	Reason of Unsuccessfulness
1	Table – Operating, Electro Hydraulic	Amnooc international medical Gmbh	No EFDA Product Certificate, Safe working load: 200Kg against at least 220 Kg
		Ceq ltd	No EFDA Product Certificate, Height adjustment 735-1065mm Against 690-940mm
		Infrinium medical inc.	No EFDA Product Certificate, Height adjustment: 740-1100mm Against 690-940mm, Lateral tilt (left/right): ±15 Against ±20 deg.
		Nanjing Mindray Bio-medical Electronics Co., Ltd.	Safe working load: 185Kg against at least 220 Kg, Table top rotation: 180° not indicated, Height adjustment: 720-1070mm Against 690-940mm
		PREMIUM MEDICAL TECHNOLOGIES FZE	No ISO 13485 product Certificate from SRA for quoted product
		GEKI Star Trading FZE	No EFDA product certificate, No ISO 13485 Certificate from SRA for quoted product Table top rotation: 180° not indicated
		Medifa Gmbh and Co. KG	No EFDA Product Certificate for quoted Model

(Handwritten signature)

ለስያ ባለቤቱ ስህተት ለማስተካከል ይህንን ሰነድ ይጠቀሙ
Make sure that this is the correct issue before use



OF for Reasons for Rejection: OF/TM/NO-17, 0, 13/09/2019