

ቀን 12 JAN 2022
Date
ቁጥር
Ref. No. EP/SA/06/543/01

To: All Participants of Tender Number ICB/EP/SA6/RDF-Routine/MI/21/21

Subject: Clarification

Reference is made to your letter dated on Dec. 23, 2021 by which you have requested clarification on the above caption tender.

We hereby clarify that the agency requested steam sterilizer with the minimum requirement of volume capacity 45L stated in the bid document. So, all the product evaluation will be done against the stated detail technical specification in the bid document.

We hope the above points clarify your request.

With Best Regards

CC:

- Deputy Director General (Inbound Logistics)
- Medical Device Tender Management Team
- Archive

EP/SA