

31 JAN 2022

ቀን
Date

ቁጥር
Ref. No.

EPISA/06/604/04

To: All participants of Tender number ICB/EPISA6/RDF, SDG & GF/MI/24/21

Subject: Clarification on tender No. ICB/EPISA6/RDF, SDG & GF/MI/24/21

Dear Sirs,

Reference is made to your letter dated on January 21 and 24, 2022 by which you have requested clarification on the above caption tender.

We hereby declare that;


Ø For item no. 7 (Laundry Machine); in our bidding document under Part-2 Statement of Requirement section -6 the requested trolley is external separate accessory made of SS 316 which is laundry trolley for wet Linens.

Ø For item no. 19 (Warmer – Radiant, Infant); in our bidding document under Statement of Requirement section we requested:

1. Double Swiveling castors wheel is coupled castor turn around appoint.
2. As the warmer is open the control unit needs to monitor skin and air /room temperature to drive radiant heater output (servo& manual).
3. Integrated timer with count up and down is for heating time of the neonate.
4. Regarding with oxygen cylinder capacity please quote as per the request.

We hope the above points clarify your request.

With Best Regards



CC:

- Deputy Director General (Inbound)
- Medical Device Tender Management Team
- Archive
EPISA

