

ቀን 16 FEB 2022.
Date
ቁጥር
Ref. No. EPISA/06/646/01

To: All participants of Tender No. ICB/EPISA6/RDF, SDG & GF/MI/24/21

Subject: Clarification on tender No. ICB/EPISA6/RDF, SDG & GF/MI/24/21

Dear Sirs,

Reference is made to your letter dated on January 21, 28 & 31, 2022 by which you have requested clarification on the above caption tender.

We hereby declare that;

Ø The final destination for all requested items is Djibouti for Sea and Addis Ababa for air.

Ø For item no.1(Tube Sealer -Automatic); in our bidding document under Part-2 Statement of Requirement section -6 the requested Tube sealer amended to :-

1. Automatic bench top type tube sealer, powered by battery with charger (220 VAC, +- 10%, 50 HZ).
2. Number of seal per full charge should be more than 1000.
3. Simultaneous Multiple tube sealing up to 4 tubes.

Ø For item no.4 (Ultrasound- Ob/Gyn); in our bidding document under Part-2 Statement of Requirement section -6 the requested Ultrasound is Color Doppler.

Ø For item no. 5 (Ultrasound-General Purpose, Color Doppler, Mobile, 4D); in our bidding document under Statement of Requirement section the requested:

1. Printer is thermal with ultrasound paper, quantity at least 100 rolls.
2. Endo cavity probe trans-vaginal (4-9 MHz) applicable for examining internal organ such as vaginal, Cervix, uterus, fallopian, ovaries rectum etc.

We hope the above points clarify your request.

With Best Regards

CC:

- Deputy Director General (Inbound Logistics)
 - Medical Device Tender Management Team
 - Archive
- EPISA

