

ቀን 19 MAY 2022
Date
ቁጥር EPISA/06/948/04
Ref. No.

To: All participants of tender No. ICB/EPISA6/RDF, SDG & GF/MI/24/21

Dear Sir/Madam,

Subject:- First Partial Tender Result:

First partial assessments of the above captioned tender are now finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Pharmaceutical Supply Agency (EPISA) Director General Office from May 19, 2022 G.C. to May 25, 2022 G.C.; 8:30AM-12:30PM and 1:30PM-5:30PM.

With regards,



Hany Mogessie
Pharmaceuticals & Medical
Supplies Procurement
Management Director
P/Director



ENC.: 02 Page

CC:

- Director General
- Deputy Director General/Inbound Logistics/
- Medical Device Tender Management Case Team
- Quantification and Market Shaping Directorate

EPISA



Company Name: የኢትዮጵያ ሎጅስቲክስ አገልግሎት
 Ethiopian Pharmaceuticals Supply Agency
 Title: First Partial Official Winner list
 Document No: OF/TM/ HO-15
 Effective date : 13/09/2019
 Rev. No.: 0
 Page 1 of 1

Tender No.: ICB/EPSA6/RDF, SDG & GF/M/24/21

Item No	Description	Unit	Qty	Winner		Currency	Unit FOB Price	Unit Installation and Training Cost in ETB	Turnkey Cost in ETB	Incoterm	Delivery Period	Remark
				Supplier	Country							
7	Laundry Machine	Each	34				Cancelled					
15	Bed - Patient, Fowler, Adult	Each	5,000				Cancelled					
19	Warmer - Radiant, Infant	Each	86				Cancelled					
17	Cylinder - Oxygen, 50L	Each	3,802	Zaman Projects Management	China	USD	169	NA	NA	FOB	60-90 Days	
23	ECT Machine	Each	13				Cancelled					
24	Cylinder - Oxygen, 30L	Each	228	Zaman Projects Management	China	USD	132	NA	NA	FOB	60-90 Days	
25	Scale - Weight, Adult, Digital	Each	2,000				Cancelled					
26	Resuscitator - Adult, Pediatric & Infant	Set	3,500	Mission Pharma A/S	China	USD	445.00	NA	NA	FOB	60-90 Days	
27	Pulse Oximeter	Each	2,000				Cancelled					

አባቶች ለዚህ ሰነድ ከመጠቀም በፊት ትክክለኛ መሆኑን ያረጋግጡ
 Make sure that this is the correct issue before use



OF for Official winner list OF/TM/HO-16, 0, 13/09/2019

Unsuccessful Bidders Under Preliminary Evaluation		
S.No	Item Name	Reason of Unsuccessfulness
17	Cylinder - Oxygen, 50L	Delivery Period 260Days for Item No. 17 Against 60-90Days
26	Resuscitator - Adult, Pediatric & Infant	Bid Bond Validity 114Days Against 148Days

Unsuccessful Bidders Under Technical Evaluation				
Item No	Item Name	Reason of Unsuccessfulness		
17	Cylinder - Oxygen, 50L	KBM Engineering Limited	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region	
			Fira MED Trading PLC	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region
			Fenway Trading	No Valid EFDA & ISO 13485 Product Certificate From SRA Region for Cylinder manufacturer
			Empire Enterprise	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region
24	Cylinder - Oxygen, 30L	KBM Engineering Limited	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region	
			Fenway Trading	No Valid EFDA & ISO 13485 Product Certificate From SRA Region for Cylinder manufacturer
			Fira MED Trading PLC	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region
			Empire Enterprise	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region
25	Scale - Weight, Adult, Digital	Apothecaries Sundries Manufacturing Co.	No Valid ISO 13485 Product Certificate from SRA region	
			Fira MED Trading PLC	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region
			Quingado Hair	No Valid EFDA Product Certificate & No Valid ISO 13485 Product Certificate from SRA region
			Apothecaries Sundries Manufacturing Co.	Compressible self-refilling ventilation bag Adult 1600ml Against 2000ml; Infant 250ml Against 300ml
26	Resuscitator - Adult, Pediatric & Infant	Fira MED Trading PLC	Compressible self-refilling ventilation bag: Adult 1630ml Against 2000ml, Ped. 680ml Against 1000ml & Oxygen reservoir bag: Adult 1630ml Against 2000ml, Ped. 680ml Against 1000ml.	
			Quingado Hair	No Valid ISO 13485 Product Certificate from SRA region
				No Valid ISO 13485 Product Certificate from SRA region

