

ቀን 22 AUG 2022
Date
ቁጥር
Ref. No. EP/SA/06/100/01


To: All Participants of Tender with Ref. No. ICB/EP/SA6/RDF-Routine/MI/21/21

Dear Sir/Madam,

Ref.:- 2nd Partial Tender Result

A second partial assessment of the above captioned tender is now finalized and please finds the attached award list. Any question related to the assessment and award list could be forwarded to Pharmaceutical Supply Agency (EPSA) Director General Office from August 22, 2022 G.C. to August 26, 2022 G.C.; 8:30AM-12:30PM and 1:30PM-5:30PM.

With regards,



ENC.: 02 Pages

CC:

- Director General
- Deputy Director General/Inbound Logistics/
- Medical Device Tender Management Case Team
- Quantification and Market Shaping Directorate

EP/SA






Company Name:		Document No: OF/TM/ HO-15
Ethiopian Pharmaceuticals Supply Agency		Effective date : 13/09/2019
Title:		Rev. No.: 0
Official Winner list		

Second Partial Tender No.: ICB/EPSA6/RDF-Routine/MI/21/21

Item No	Description	Unit	Qty	Winner		Currency	Unit FOB Price	Unit Installation and Training Cost in ETB	Turnkey Cost in ETB	Incoterm	Delivery Period	Remark
				Supplier	Manufacturer Country							
21	Set-Episiotomy/ Perineal/Vaginal Cervical	Set	172	Beeveejay Ltd	Germany	USD	511.50	NA	NA	FOB	60-90 Days	
30	Sphygmomanometer-Digital, Adult	Each	2951	Hawann International	Japan	USD	262.00	NA	NA	FOB	60-90 Days	
31	Sphygmomanometer-Digital, Ped	Each	2951	Hawann International	Japan	USD	262.00	NA	NA	FOB	60-90 Days	

Note: Reason of Non-Compliance for the Rest Bidders: Price Incompetency

አባቶች ለዚህ ስነ-ምግባር ለፊት ትክክለኛ መሆኑን ያረጋግጡ
Make sure that this is the correct issue before use




OF for Official winner list OF/TM/HO-16, 0, 13/09/2019



Company Name: Ethiopian Pharmaceuticals Supply Agency
 የኢትዮጵያ ምድባሃት አቅርቦት አገልግሎት
 Title:
 Unsuccessful Bidders on Technical Evaluation

Second Partial Tender No.: ICB/EPPSA6/RDF-Routine/MI/21/21

Item No	Item Description	Unit	Qty	Bidder	Compliance	Reason of Non-Compliance	Remark
21	Set-Episiotomy/ Perineal/Vaginal Cervical	Set	172	Boston Medical FZE	Not Comply	<input type="checkbox"/> No Valid EFDA Product Certificate	
				Durable Hospital Supplies	Not Comply	<input type="checkbox"/> No Valid EFDA Product Certificate <input type="checkbox"/> No Valid ISO 13485 Product Specific Certificate from SRA Region	
				Avishkar	Not Comply	<input type="checkbox"/> Expired EFDA Product Certificate <input type="checkbox"/> No Valid ISO 13485 Product Certificate from SRA Region	
				Apothecaries	Not Comply	<input type="checkbox"/> No Valid ISO 13485 Product specific Certificate from SRA Region	
30	Sphygmomanometer-Digital, Adult	Each	2951	Qingdao Haier Biomedical	Not Comply	<input type="checkbox"/> Heart Rate: 40-200 bpm Against 30-250 bpm	
				Jaingsu Yuyue	Not Comply	<input type="checkbox"/> Heart Rate: 40-200 bpm Against 30-250 bpm	
				Avishkar	Not Comply	<input type="checkbox"/> Expired EFDA Product Certificate <input type="checkbox"/> No Valid ISO 13485 Product Certificate from SRA Region	
				Apothecaries	Not Comply	<input type="checkbox"/> No Valid USFDA/CE Product Certificate from SRA Region	
31	Sphygmomanometer-Digital, Ped.	Each	2951	Qingdao Haier Biomedical	Not Comply	<input type="checkbox"/> Heart Rate: 40-200 bpm Against 30-250 bpm	
				Avishkar	Not Comply	<input type="checkbox"/> Expired EFDA Product Certificate <input type="checkbox"/> No Valid ISO 13485 Product Certificate from SRA Region	
				Apothecaries	Not Comply	<input type="checkbox"/> No Valid USFDA/CE Product Certificate from SRA Region	



[Handwritten signature]