

ቀን 18 AUG 2023
Date
ቁጥር
Ref. No. EPSSA/06/149/04

To; - All participants of tender number ICB/EPSS6/GF-CHEM/MS/LS/09/23

Addis Ababa

Subject; -Amendment and Response for clarification request

Dear Sir/Madam

Thank you for your interest in participating on the above mentioned tender. You have requested clarification on the above mentioned tender; the clarification response is stated as follows:-

- 1 .For item No. 2 (Blood collecting tube-PL K2EDTA, 4ml) the item description is Blood Collecting Tube - PL K2 EDTA, 4 ml with vacuum
2. The item description and specification for item No.3 (Vacutainer Needle - Multiple Blood Sampling, 21G -), 4 (Plastic dropper - Non sterile size hold 7ml of liquid with graduated marking in 0.5ml increments) and 5 (Capillary Tube, Ethylene Diamine Tetra Acetic Acid (EDTA), 75mm Length, 1.1-1.2 Internal diameter, 1.5-1.6mm Outside Diameter) are as per the bid document.

The following points in the bid document are also amended as follows:

- Each local agent shall only be authorized by one bidder for the participation of the bid if local agent authorized by more than one bidder, the bidders shall be disqualified.
- Item no 1.Clothianidin 50%W/W – 150, the awarded Bidder shall carry out pre- shipment testing and the official result shall be delivered to EPSS and the cost of the testing will be covered by the bidder.

N.B: others terms and conditions of the bid will remain the same.

With best regards

CC:

- ✓ Deputy Director General (inbound logistic)
- ✓ Chemical and medical supplies team
- ✓ Archive

EPSS

