

**TO: All Participants of Tender no ICB/EPSS6/RDF/MD/ORHB/16/23**

Dear Sirs,

**Ref: Tender Result**

An assessment of the above captioned tender is now finalized and please find the attached award list.

Any question related to the assessment could be forwarded to Ethiopian Pharmaceutical Supply Service (EPSS) Director General Office from January 10, 2024, to January 16, 2024 8:30AM -12 :30 PM and 1:30 PM- 5:30 PM.

With Best Regards,

  
  
Tibebu Hailu Mogesse  
Pharmaceuticals & Medical  
Supplies Procurement Tender  
Management Directorate  
PI/Director

ENC: 21 page

CC

- Director General
- Deputy Director General /Inbound Logistic/
- Medical Device Tender Management Team
- Quantification & Market Shipping Directorate
- Archive

EPSS





Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.: ICB/EPPSS6/RDF/MD/ORHB/16/23

Document No: OE/TM/HO-02

Effective date : 21/01/2022


Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
1	HIGH RESOLUTION ULTRASOUND, COLO UR DOPPLER	Each	15	Fujifilm Middle East FZE	Not Comply	Active Ports: 4 Against The requested 5 active ports. Dynamic Range 272 dB Against to 280 dB/or more	
				Vinno Technology(Suzhou)Co.,Ltd	Not Comply	Convex Array Transducer with frequency range for Abdominal, Ob/Gyn applications are 2-8.5Mhz against to 1 to 7 Mhz, Intracaviary Transducer frequency range is 5-11 MHz against to 3 to 15 Mhz, Volume 4D Convex Transducer frequency range is 3-6 MHz against to 2-7 MHz, Linear Transducer frequency range is 6-16 MHz against to 4-16 MHz, Operating System is window against to Linux system	
				Adonai Trading FZE	Not Comply	Convex Array Transducer with frequency range for Abdominal, Ob/Gyn applications are 2-5.4Mhz against to 1 to 7 Mhz, Intracaviary Transducer frequency range is 3-10 MHz against to 3 to 15 Mhz, Volume 4D Convex Transducer frequency range is 2-6 MHz against to 2-7 MHz, Linear Transducer frequency range is 4-13.3 MHz against to 4-16 MHz, Operating System is window against to Linux system	

OE for Technical Evaluation, OE/TM/HO-01, 21/01/2022

አቅጣጫ ስራው ስለተገኘው በረዕይ  
Make sure that this is the correct





**Company Name:** የኢትዮጵያ መድኃኒት አቅጣጫ አገልግሎት  
**Ethiopian Pharmaceuticals Supply Agency**  
**Title:** Unsuccessful Bidder Under Technical Evaluation  
**Document No: OF/TM/ HO-02**  
**Effective date : 21/01/2022**  
**Rev. No.: 01**  
**Note:- Reason of rejection: Offer against the request**

**Medical Device Tender with Ref. No.: ICB/EPSS6/RDF/MD/ORHB/16/23**

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
				Samsung Medison Co., Ltd	Not Comply	Operating system is window against to linux system , sample gate size is 0.5-2.5 mm against to 1-40 mm, archive storage is external CD/DVD against to inbuilt storage system, Intracavitary Transducer frequency range is 2-11 MHz against to 3-15 Mhz	
				Quality medical FZC	Not Comply	Convex Array Transducer with frequency range for Abdominal Ob/Gyn applications are 2-6MHz against to 1 to 7 Mhz, Intracavitary Transducer frequency range is 4-9 MHz against to 3 to 15 Mhz, Volume 4D Convex Transducer frequency range is 2-4 MHz against to 2-7 MHz, Linear Transducer frequency range is 4-12 MHz against to 4-16 MHz, factory level training is not provided.	
				Wuhan Zoncare Biomedical Electronics Co., Ltd	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	EFDA Model specific product Certificate not submitted and Submitted SRA Product Certificate is expired	
				Elismed Healthcare Solutions	Not Comply	EFDA Model specific product Certificate not submitted and No Valid Product Certificate from SRA Region for the Quoted Manufacturer site	

ረባረባ በዚህ ስነ፡ ኮምፕላዥንት ስር፡  
 Make sure that this is the

**OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022**





**Company Name:** የኢትዮጵያ ሙድሃሪት አኅሊብ ኤጀንሲ  
**Ethiopian Pharmaceuticals Supply Agency**  
**Title:** Unsuccessful Bidder Under Technical Evaluation  
**Document No:** OF/TM/ HO-02  
**Effective date :** 21/01/2022  
**Rev. No.:** 01  
**Note:- Reason of rejection: Offer against the request**

**Medical Device Tender with Ref. No.: ICB/EPPSS6/RDF/MD/ORHB/16/23**

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
2	MOBILE X-RAY MACHINE DIGITAL	Each	10	Fujifilm Middle East FZE	Not Comply	2nd priority	Valid model specific EFDA certificate for the Quoted Manufacturer site not submitted and Valid Product Certificate from SRA Region for the Quoted Manufacturer site is not submitted
				Quality Medical FZC	Not Comply	Submitted SRA Product Certificate is Expired	
				Boston Medical FZE	Not Comply	2nd priority	
				ABI Business Group drugs & Medical Equipment importer Plc	Not Comply	2nd priority	
3	DIGITAL X-RAY MACHINE	Each	15	Elismed Healthcare Solutions	Not Comply	Valid model specific EFDA certificate for the Quoted Manufacturer site not submitted and Valid Product Certificate from SRA Region for the Quoted Manufacturer site not submitted	
				Fujifilm Middle East FZE	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	2nd priority	
				ABI Business Group drugs & Medical Equipment importer Plc	Not Comply	2nd priority	
5	RADIANANT WARMER	Each	50	Quality Medical FZC	Not Comply	Submitted SRA Product certificate is Expired	
				Elismed Healthcare Solutions	Not Comply	Valid model specific EFDA certificate for the Quoted Manufacturer site not submitted and Valid Product Certificate from SRA Region for the Quoted Manufacturer site not submitted	
				Jiangsu Saikang Co.,ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Fenway Trading Co. Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	Valid Product Certificate from SRA Region for Specific Manufacturing Site not submitted	

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

አባባዎ በዚህ ላይ ከተሞላቀው በኋላ ትክክለኛ ይህን ደብዳቤ ያዘጋጁ

Handwritten signatures and official stamp of Ethiopian Pharmaceutical Supply Service.



Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Medical Device Tender with Ref. No.:ICB/EPSS6/RDF/MD/ORHB/16/23


Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
6	OXYGEN CONCENTRATOR	Each	50	BPL Medical Technologies Pvt. Ltd	Not Comply	2nd priority	
7	ANESTHESIA MACHINE WITH INTEGRATED VENTILATOR AND PATIENT MONITOR	Each	50	Adonai Trading FZE	Not Comply	Provide 3 column-flow-meters against to 6 column-flow-meters	
				Jiangsu Saikang Co., Ltd	Not Comply	Provide 3 column-flow-meters against to 6 column-flow-meters	
				Rise Global GmbH	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Fenway Trading Co., Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	
				BPL Medical Technologies Pvt. Ltd	Not Comply	2nd priority	
				Adonai Trading FZE	Not Comply	Submitted product certificate from SRA region is not valid	
11	STEAM STERILIZER	Each	30	BGI Health (HK) Co.Ltd	Not Comply	2nd priority	
				Unique Enterprise Ltd	Not Comply	2nd priority	
				Care Pharma FZ LLC	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region is not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	Valid Product Certificate from SRA Region-not submitted	

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

ላኪዎ በዚህ ስድስት ከመጠቀም በኋላ ማረጋገጥ ይደረጋል

*(Handwritten signature)*

*(Handwritten signature)*



Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት  
Ethiopian Pharmaceuticals Supply Agency

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Title:

Unsuccessful Bidder Under Technical Evaluation


Note:- Reason of rejection: Offer against the request

Medical Device Tender with Ref. No.: ICB/EPSS6/RDF/MD/ORHB/16/23

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
13	STEAM STERILIZER, 40 LITER	Each	50	Adonai Trading FZE	Not Comply	Submitted product certificate from SRA region is not valid	
				Care Pharma FZ LLC	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				BGI Health (HK) Co.,Ltd	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
20	STRETCHERS	Each	50	Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
22	GASTROINTESTINAL INSTRUMENT SET	Set	50	Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	

OF for Technical Evaluation: OF/TM/HO-01, 21/01/2022

አባቶ ባህሪ ለገጽ ከመጠቀም በፊት  
Make sure that this is the correct





Company Name:

የኢትዮጵያ ሕክምና ስጦታ አገልግሎት አ.የ.አ.የ.አ.  
Ethiopian Pharmaceuticals Supply Agency

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Title: Unsuccessful Bidder Under Technical Evaluation

Note:- Reason of rejection: Offer against the request

Medical Device Tender with Ref. No.:ICB/EPSS6/RDF/MD/ORHB/16/23

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
23	ABDOMINAL PERINEAL RESECTION SET	Set	50	Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	
25	VAGINAL HYSTERECTOMY SET	Set	50	Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	
26	ABDOMINAL GYNECOLOGICAL INSTRUMENT SET	Set	50	Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

አስከፊ ሰዓዊ ስድስት ስምንት ሰዓት  
Make sure that this is the

*(Handwritten signatures and stamps)*



Company Name: የኢትዮጵያ ሕክምና ምድሃነት አቅጣጫ አገልግሎት  
 Ethiopian Pharmaceuticals Supply Agency  
 Title: Unsuccessful Bidder Under Technical Evaluation  
 Document No: OF/TM/ HO-02  
 Effective date : 21/01/2022  
 Rev. No.: 01  
 Note:- Reason of rejection: Offer against the request

Medical Device Tender with Ref. No.: ICB/EPSS6/RDF/MD/ORHB/16/23

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
27	DIAPHRAGMATIC HERNIA REPAIR SET	Set	50	Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	
				Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
31	LAMINECTOMY SET	Set	50	Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	
				Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
33	HAEMORRHOIDECTOMY SET	Set	50	Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	2nd priority	

አስከፊ በዚህ ስነ-ምግባር ላይ ማረጋገጫ ይደረጋል  
 Make sure that this is the

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022







Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Title:


Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.: ICB/EPSS6/RDF/MD/ORHB/16/23

Item No.	Description	Quantity	Manufacturer	Compliance Status	Justification for Non-Compliance	Remarks
34	URETER DILATION SET	Set 50	Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
			Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
			Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
			Boston Medical FZE	Not Comply	2nd priority	
35	CAESAREAN SECTION SET	Set 50	Beevejay Ltd	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
			Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
			Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
			Boston Medical FZE	Not Comply	2nd priority	
42	SPHYGMOMANOMET ER ADULT,DIGITAL	Each 500	Rise Global GmbH	Not Comply	Works by Dry Battery against Rechargeable battery	
			BPL Medical Technologies Pvt. Ltd	Not Comply	Devices not read MAP, Heart Rate reading is 40 - 200bpm against 30-250 bpm, Mode of operation only support Adult against to the request for Adult, Pediatric and Infant, External AA rechargeable battery against internal rechargeable battery.	

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

አስከፊ በዚህ ላይ ከመጠቀም በፊት  
Make sure that this is the





Company Name:

የኢትዮጵያ ሕክምና ምርት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Document No: OF/TM/HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Medical Device Tender with Ref. No.:ICB/EPSS6/RDF/MD/ORHB/16/23

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
43	PATIENT BED AT LEAST THREE SECTIONS WITH MATTRESS	Each	300	Boston Medical FZE	Not Comply	Two section against to Three Section	
				Beeveejay Ltd	Not Comply	2nd priority	
				ABI Business Group drugs & Medical Equipment importer Plc	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Sevato Pharma	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
44	PEDIATRIC PATIENT BED WITH MATTRESS	Each	200	Jiangsu Saikang Co.,ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate	
				Sevato Pharma	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	

አባባዎ በዚህ ሰነድ ከመጠየቅዎ በኋላ ትክክል ይሁን ያረጋግጡ

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

*(Handwritten signatures and stamps)*



Company Name:

የኢትዮጵያ የድምጻዎች አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.:CEB/EPSS6/RDF/MD/ORHHB/16/23

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection:Offer against the request

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
45	REFRIGRATOR >300LITER	Each	50	Care Pharma FZ LLC	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Fenway Trading	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Unique Enterprises Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
50	SLIT LAMP/AUTOMATIC	Each	20	Beevejay Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	2nd priority	
59	INFANT INCUBATOR	Each	50	Boston Medical FZE	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	

አብዮታ በዚህ ላይ ከመጠቀም በፊት  
Make sure that this is the correct

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022





Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.: ICB/EPSS6/RDE/MD/ORHHB/16/23

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
60	COAGULATION MACHINE	Each	4	Roche Diagnostics	Not Comply	Valid Product Certificate from SRA Region not submitted	
				BGI Health (HK) Co.,Ltd	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Boston Medical FZE	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	
				Beevejay Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
64	SET - AMPUTATION INSTRUMENT	Set	10	Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Care Pharma FZ LLC	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	2nd priority	
67	LARGE BONE POWER DRILL WITH SAW SET	Set	10	Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevato Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	EFDA product Certificate and Submitted SRA-Region Product certificate is Expired	

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

አባኮሞ ቢህን ስነ-ምግባር ስራ ለማረጋገጥ ይህን ደብዳቤ ይሙሉ እና ይጽፉ

Make sure that this is the correct

*(Handwritten signatures)*





Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.:ICB/EPSS6/RDF/MDD/ORHHB/16/23

Document No: O/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
68	SET - INSTRUMENT, RUSH NAIL	Set	10	Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevato Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	EFDA product Certificate and Submitted SRA Region Product certificate is Expired	
69	SET - INSTRUMENT,CANULAT ED SCREW 7MM	Set	10	Jiangsu Saikang Co.,Jtd	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevato Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	EFDA product Certificate not submitted and Submitted SRA Region Product certificate is Expired	
				Rise Global GmbH	Not Comply	Reduction Forceps Pointed 203.2mm against Reduction Forceps Pointed 210mm, Reduction Forceps Pointed 254mm against	
70	SET - INSTRUMENT, LARGE FRAGMENT 4.5 NON LOCKING	Set	10	Jiangsu Saikang Co.,Jtd	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevato Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Beardsell Limited	Not Comply	Valid EFDA product Certificate not submitted and Submitted SRA Product certificate is Expired	

አብዮት ስለሆነ ይህን ደብዳቤ በሌላ ገጽ ይጻፉ  
Make sure that this is the correct issue

OF for Technical Evaluation: O/TM/HO-01, 21/01/2022





Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.:ICB/EPSS6/RD/FMD/ORHHB/16/23

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
71	SET - INSTRUMENT, SMALL FRAGMENT, 3.5MM NON-LOCKING	Set	10	Rise Global GmbH	Not Comply	Self-Centering Bone Holding Forceps 152.4mm against Self-Centering Bone	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevato Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,ld	Not Comply	2nd priority	
				Beardsell Limited	Not Comply	EFDA product Certificate not submitted and Submitted SRA Product certificate is Expired	
				Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
72	SET - FIXATOR, EXTERNAL, LARGE (SYNTHESES TYPE)	Set	10	Sevato Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,ld	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	2nd priority	

OF for Technical Evaluation-OF/TM/HO-01, 21/01/2022

አባቶ ለዚህ ሰነድ ከመጠየቅ ለፈታ ትክክል ማረጋገጥ ይጠበቃልልኝ  
Make sure that this is the correct





Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Medical Device Tender with Ref. No.: ICB/EPSS6/RDF/MD/ORHB/16/23

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Note:- Reason of rejection: Offer against the request

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
73	SET - INSTRUMENT, HCS/DHS	Set	10	Rise Global GmbH	Not Comply	Not gouted for Tap 6.5mm, Drill and tap sleeve 6.5/4.5mm, Drill and tap sleeve 4.5/3.5mm & 4.5/3.5mm, Drill bit 4.5mm, T Wrench and Long coupling sleeve for removal	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevatio Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
74	SET - PELVIC INSTRUMENT	Set	10	Sevatio Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Rise Global GmbH	Not Comply	1x Pelvic Reduction Forceps with ball Tip Oblique 190mm against 2x Reduction forceps oblique	
				Boston Medical FZE	Not Comply,	2nd priority	
				Jiangsu Saikang Co., Ltd	Not Comply	2nd priority	

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022

Make sure that this is the





**Company Name:** የኢትዮጵያ ሕክምና ስጪት አገልግሎት አ.ኃ.የ/አ  
**Ethiopian Pharmaceuticals Supply Agency**  
**Title:** Unsuccessful Bidder Under Technical Evaluation  
**Document No:** OF/TM/ HO-02  
**Effective date :** 21/01/2022  
**Rev. No.:** 01

**Medical Device Tender with Ref. No.:(CB/EPSS6/RDP/MD/ORH/16/73)**

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
75	SET - INSTRUMENTS, SMALL FRAGMENT LOCKING.(3.5)	Set	10	Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevatto Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Rise Global GmbH	Not Comply	1xdrill sleeve for locking screw 2.7mm against 3xlocking guide	
				Jiangsu Saikang Co.,Ltd	Not Comply	2nd priority	
				Beardsell Limited	Not Comply	EFDA product Certificate not submitted and Submitted SRA Product certificate is Expired	
				Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Sevatto Pharma	Not Comply	EFDA product Certificate and Valid Product Certificate from SRA Region not submitted	
				Jiangsu Saikang Co.,Ltd	Not Comply	2nd priority	
				Beardsell Limited	Not Comply	EFDA product Certificate not submitted and Submitted SRA Region Product certificate is Expired	
76	SET - INSTRUMENT, LARGE FRAGMENT LOCKING	Set	10	Boston Medical FZE	Not Comply	2nd priority	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Rise Global GmbH	Not Comply	Wire Sleeve 3.2mm X 150mm against Wire Sleeve Ø 8.0 X 1.8mm, Wire Sleeve 4.0mm X 150mm against Wire Sleeve Ø 8.0 X 4mm	
				Boston Medical FZE	Not Comply	2nd priority	
77	SET - INSTRUMENT SYSTEM, TIBIA FEMURE	Set	10	Beardsell Limited	Not Comply	EFDA product Certificate not submitted and Submitted SRA Product certificate is Expired	
				Boston Medical FZE	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	2nd priority	

አስቀድሞ በዚህ ስጪት ስም ለማግኘት ባልቻሉ ገንዘብ ላይ  
 Make sure that this is the correct

ማስረጃ ለቴክኒካል ማጠቃለያ ማስፈሰስ ማድረግ  
 Offer for Technical Evaluation, OF/TM/HO-01, 21/01/2022





Company Name:

የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት

Ethiopian Pharmaceuticals Supply Agency

Title:

Unsuccessful Bidder Under Technical Evaluation

Note:- Reason of rejection: Offer against the request

Medical Device Tender with Ref. No.: ICB/EPPSS6/RDF/MD/ORH/B/16/23

Document No: OF/TM/ HO-02

Effective date : 21/01/2022

Rev. No.: 01

Item No	Item Name	Unit	Qty	Bidder	Compliance	Reason of non-compliance	Remark
78	SET - INSTRUMENT PENNA II	Set	10	Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Rise Global GmbH	Not Comply	Protection Sleeve 17mm against Trocar 3mm, PF Flexible Reamer 16.2mm for Nail against Proximal Reamer 15mm.	
				Jiangsu Saikang Co.,ltd	Not Comply	2nd priority	
				Beardsell Limited	Not Comply	EFDA product Certificate not submitted and Submitted SRA Region Product certificate is Expired	
79	BASIC ORTHOPEDIC SET	Set	10	Boston Medical FZE	Not Comply	2nd priority	
				Beevejay Ltd	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Narang Medical Limited	Not Comply	Valid Product Certificate from SRA Region not submitted	
				Rise Global GmbH	Not Comply	Not quote for Hohmann retractor with width of - 12mm, 16mm, 24mm	
80	CPAP MACHINE	Each	4	Boston Medical FZE	Not Comply	2nd priority	
				Boston Medical FZE	Not Comply	EFDA Model specific product Certificate and Valid Product Certificate from SRA Region not submitted	

አባኮታ ባለቤት ስያሜ ከመጠየቅ ስራ ላይ  
Make sure that this is the correct

OF for Technical Evaluation, OF/TM/HO-01, 21/01/2022



Tender No.: ICB/EPSS6/RDF/MD/ORRH/16/23

Company Name: ገብጽ ስፔሻላይዥን ፊርማ ስፔሻላይዥን ፊርማ  
Ethiopian Pharmaceuticals Supply Agency  
Title: Official Winner List

Document No: OFTM/HO-15  
Effective date : 21/01/2022  
Exchange Date: Oct. 02, 2023  
Rev. No.: 01  
Page 1 of 1

Item No	Description	Unit	Qty	Bidder	Manufacturer	Currency	Unit FOB Price	Unit Installation and Training Cost in ETB	Turnkey cost in ETB	INCOTERM	Delivery Period (Day)	Result	Remark
1	HIGH RESOLUTION ULTRASOUND, COLOUR DOPPLER	Each	15	Spot Medical Equipment FZE	Esate S.P.A	EUR	46,000.00	35,000.00		FOB	60 Days	Winner	
2	MOBILE X-RAY MACHINE DIGITAL	Each	10	Shenzen Browiner Co. Ltd	Shenzen Browiner Tech Co. Ltd	USD	66,450.00	44,400.00	55,296.00	FOB	90 Days	Winner	
3	DIGITAL X-RAY MACHINE	Each	15	Shenzen Browiner Co. Ltd	Shenzen Browiner Tech co. Ltd	USD	73,320.00	30,000.00	110,594.00	FOB	90 Days	Winner	
4	PATIENT MONITOR	Each	50	Beardsell Limited	Nice Neotech Medical System Pvt Ltd	USD	7,288.50	3,076.14		FOB	90 Days	Winner	
5	RADIANT WARMER	Each	50	Rise Global GmbH	Contec Medical System Co. Ltd	USD	312.00			FOB	60-90 Days	Winner	
6	OXYGEN CONCENTRATOR	Each	50	Spot Medical Equipment FZE	Heyer Medical AG	EUR	18,000.00	45,000.00		FOB	60 Days	Winner	
7	ANESTHESIA MACHINE WITH INTEGRATED VENTILATOR AND PATIENT MONITOR	Each	50	ABI Business Group Drugs and medical equipment importer PLC	Shirva Medical Instrument co. Ltd	ETB	2,661,349.60	354,800.00		COD	90 Days	Winner	
8	OPERATION LIGHT CEILING TYPE	Each	50	ABI Business Group Drugs and medical equipment importer PLC	Shirva Medical Instrument co. Ltd	ETB	411,310.12	78,758.00		COD	90 Days	Winner	
9	MOBILE OPERATION LIGHT	Each	50							Cancelled			
10	ELECTRO CARDIOGRAPH (ECG) MACHINE	Each	50							Cancelled			
11	STEAM STERILIZER	Each	30							Cancelled			
12	STEAM STERILIZER >80L	Each	50							Cancelled			
13	STEAM STERILIZER, 40 LITER	Each	50							Cancelled			
14	OPERATION TABLE, MULTI SECTION, ELECTRO -HYDRAULIC	Each	50							Cancelled			
15	SURGICAL SUCTION MACHINE, ELEC, 2 BOTTLE	Each	50							Cancelled			
16	VACUUM EXTRACTOR, MANUAL	Each	20							Cancelled			
17	VACUUM EXTRACTOR, ELECTRICAL	Each	20							Cancelled			
18	ELECTROSURGICAL UNIT	Each	50							Cancelled			
19	DELIVERY COACH	Each	50							Cancelled			
20	STRETCHERS	Each	50	Boston Medical FZE	Better Medical Technologies Co., Ltd	USD	388.41	226.92		FOB	60-90 Days	Winner	
21	EXAMINATION COACH	Each	50							Cancelled			
22	GASTROINTESTINAL INSTRUMENT SET	Set	50	Rise Global GmbH	Casis International	USD	514.22			FOB	60-90 Days	Winner	
23	ABDOMINAL PERINEAL RESECTION SET	Set	50	Rise Global GmbH	Casis International	USD	236.64			FOB	60-90 Days	Winner	
24	MAJOR RECTAL INSTRUMENT SET	Set	50	Rise Global GmbH	Casis International	USD	160.51			FOB	60-90 Days	Winner	
25	VAGINAL HYSTERECTOMY SET	Set	50	Rise Global GmbH	Casis International	USD	160.51			FOB	60-90 Days	Winner	

አባቶች ስለዚህ ዝርዝር ማረጋገጫ ላይ ምልክት ያድርጉ።  
Make sure that this is the correct issue before use

OR for Financial Evaluation, OFTM/HO-14 0, 21/01/2022

*(Handwritten signatures)*



Item No	Description	Unit	Qty	Bidder	Manufacturer	Currency	Unit FOB Price	Unit Installation and Training Cost in ETB	Turkey cost in ETB	INCOTERM	Delivery Period (Day)	Result	Remark
26	ABDOMINAL GYNECOLOGICAL INSTRUMENT SET	Set	50	Rise Global GmbH	Casis International	USD	149.76			FOB	60-90Days	Winner	
27	DIAPHRAGMATIC HERNIA REPAIR SET	Set	50	Rise Global GmbH	Casis International	USD	384.93			FOB	60-90Days	Winner	
28	LARYNGECTOMY SET	Set	50									Cancelled	
29	CRANIOTOMY SET	Set	50									Cancelled	
30	PROSTATECTOMY SET	Set	50									Cancelled	
31	LAMINECTOMY SET	Set	50	Rise Global GmbH	Casis International	USD	743.43			FOB	60-90Days	Winner	
32	CHOLECYSTECTOMY SET	Set	50									Cancelled	
33	HAEMORRHOIDECTOMY SET	Set	50	Rise Global GmbH	Casis International	USD	239.60			FOB	60-90Days	Winner	
34	URETER DILATION SET	Set	50	Rise Global GmbH	Casis International	USD	204.44			FOB	60-90Days	Winner	
35	CASABEAN SECTION SET	Set	50	Rise Global GmbH	Casis International	USD	335.69			FOB	60-90Days	Winner	
36	INCISION & DRAINAGE SET	Set	50									Cancelled	
37	OTOSCOPE, HANDHELD SET	Set	50									Cancelled	
38	LARYNGOSCOPE SET	Set	50									Cancelled	
39	EXAMINATION LIGHT	Set	50									Cancelled	
40	HOSPITAL LAUNDRY MACHINE WITH IRONING	Each	15									Cancelled	
41	LAPAROTOMY SET	Set	50									Cancelled	
42	SPRYGOMANOMETER	Each	500	Spot Medical Equipment FZE	A & D Medical	EUR	280.00			FOB	60 Days	Winner	
43	PATIENT BED AT LEAST THREE SECTIONS WITH MATTRESS	Each	300	Adonal Trading FZE	Meyasa Tibbi Chizari AS	USD	500.00			FOB	90 Days	Winner	
44	PEDIATRIC PATIENT BED WITH MATTRESS	Each	200	Boston Medical FZE	Better Medical Technology Co. Ltd	USD	288.18		168.36	FOB	60-90Days	Winner	
45	REFRIGERATOR >=300LITER	Each	50	Spot Medical Equipment FZE	Helmer Scientific	EUR	6,650.00		35,000.00	FOB	60 Days	Winner	
46	DEBRIDATOR	Each	30									Cancelled	
47	INFUSION PUMP	Each	50									Cancelled	
48	CENTRIFUGE	Each	30									Cancelled	
49	MICROSCOPE,LED	Each	50									Cancelled	
50	SLIT LAMP AUTOMATIC	Each	20	Rise Global GmbH	Chongqing Kanghuakui Ming science Technology co. Ltd	USD	2,520.00			FOB	60-90 Days	Winner	
51	OPHTHALMOSCOPES, DIRECT	Each	40									Cancelled	
52	VISUAL SIGN	Each	20									Cancelled	
53	OPERATING MICROSCOPE	Each	2									Cancelled	
54	CATARACT SET	Set	10									Cancelled	
55	ASCAN BIOMETRY	Each	5									Cancelled	
56	CARE TONOMETRY	Each	15									Cancelled	
57	DENTAL UNIT ADVANCE	Each	10									Cancelled	
58	PHOTOTHERAPY	Each	50									Cancelled	
59	INFANT INCUBATOR	Each	50	Fenway Trading Co. Ltd	Bistos Co. Ltd	USD	3,520.00			FOB	60-90 Days	Winner	
60	COAGULATION MACHINE	Each	4	Unique Enterprises Ltd	Diagnostica Sego S.A.S	USD	33,230.77			FOB	60-90 Days	Winner	

Make sure that this is the correct issue before use




OR for Financial Evaluation, OF/TM/FO-14 0, 21/01/2022

Tender No: IC/EPSS6/RDF/MD/ORRH/16/23

Company Name: የኢትዮጵያ ጤና ሚኒስቴር  
Ethiopian Pharmaceuticals Supply Agency  
Title: Official Winner List  
Exchange Date: Oct 02, 2023  
Rev. No.: 01  
Delivery Period (Day):

Item No	Description	Unit	Qty	Bidder	Manufacturer	Currency	Unit FOB Price	Unit Installation and Training Cost in ETB	Turnkey cost in ETB	INCOTERM	Delivery Period (Day)	Result	Remark
61	FLUORESCENCE MICROSCOPE	Each	4					Cancelled					
62	MICROTOME (TISSUE SLICER) DIGITAL	Each	4					Cancelled					
63	TISSUE PROCESSING MACHINE (AUTOMATED)	Each	4					Cancelled					
64	SET - AMPUTATION INSTRUMENT	Set	10	Rise Global GmbH	Nebula Surgical Pvt Ltd	USD	420.00			FOB	60-90 Days	Winner	
65	HUMBLY KNIFE AND BLADES	Each	10					Cancelled					
66	TOURNOUET-PNEUMATIC LARGE BONE POWER DRILL WITH SAW SET	Set	10					Cancelled					
67	SET - INSTRUMENT, RUSH NAIL, SCREW 7MM	Set	10	Rise Global GmbH	Nebula Surgical Pvt Ltd	USD	2,765.00			FOB	60-90 Days	Winner	
68	SET - INSTRUMENT, RUSH NAIL, SCREW 7MM	Set	10	Rise Global GmbH	Nebula Surgical Pvt Ltd	USD	280.00			FOB	60-90 Days	Winner	
69	SET - INSTRUMENT, CANDLATED	Set	10							FOB	60-90 Days	Winner	
70	SET - INSTRUMENT, LARGE FRAGMENT, 4.3 NON LOCKING	Set	10							FOB	60-90Days	Winner	
71	SET - INSTRUMENT, SMALL FRAGMENT, 3.5MM NON LOCKING	Set	10							FOB	60-90 Days	Winner	
72	SET - FIXATOR, EXTERNAL, LARGE (SYNTHES TYPE)	Set	10	Rise Global GmbH	Nebula Surgical Pvt Ltd	USD	450.00			FOB	60-90 Days	Winner	
73	SET - INSTRUMENT, HCS/DHS	Set	10							FOB	60-90Days	Winner	
74	SET - PELVIC INSTRUMENT	Set	10							FOB	60-90Days	Winner	
75	SET - INSTRUMENTS, SMALL FRAGMENT LOCKING (3.5)	Set	10							FOB	60-90Days	Winner	
76	SET - INSTRUMENT, LARGE FRAGMENT LOCKING	Set	10	Rise Global GmbH	Nebula Surgical Pvt Ltd	USD	1,134.00			FOB	60-90Days	Winner	
77	SET - INSTRUMENT SYSTEM, TIBIA FEMURE	Set	10							FOB	60-90Days	Winner	
78	SET - INSTRUMENT PENAL II	Set	10							FOB	60-90Days	Winner	
79	Basic Orthopedic Set	Set	10							FOB	60-90Days	Winner	
80	CPAP MACHINE	Each	4	Ambulance(Shenzhen) Tech Co., Ltd	Ambulance(Shenzhen) Tech Co., Ltd	USD	4,150.00		12,000.00	FOB	40 Days	Winner	

አስፈላጊ ማረጋገጫ ለውጭ አገር ማስተካከያ ማድረግ  
Make sure that this is the correct issue before use

OF for Financial Evaluation, OF/TM/HO-14 0, 21/01/2022



Handwritten signatures in blue ink.



Company Name: የኢትዮጵያ መድኃኒት አቅርቦት አገልግሎት  
 Ethiopian Pharmaceuticals Supply Agency  
 Title: Unsuccessful Bidder Under preliminary Evaluation  
 Document No: OF/TM/HO-02  
 Effective date : 21/01/2022  
 Rev. No.: 01

Item No	Bidder	Compliance	Reason of non-compliance	Remark
1	Akhyar Surgical Professionals	Not Comply	One Local Agent with Three Bidder	
2	Al-Muthahawerah	Not Comply	One Local Agent with Two Bidder	
3	B Braun meismungen AG	Not Comply	One Local Agent with Three Bidder	
4	Barnai and Hetiao Hongkong Limited	Not Comply	Bid Security Validity period: 120 Days Against 148 Days	
5	Beijing Aeommed Co., Ltd	Not Comply	One Local Agent with Seven Bidder	
6	DRGEM	Not Comply	Delivery Period: 150-190 Days Against 60-90 Days	
7	Farem Ltd	Not Comply	One Local Agent with Seven Bidder	
8	Gokler A.S	Not Comply	One Local Agent with Five Bidder	
9	DragerWerk AG&Corporate	Not Comply	Delivery Period: 90-120 Days Against 60-90 Days	
10	Guangdong Biolight Meditech	Not Comply	One Local Agent with Seven Bidder	
11	Hebei Pukang Medical Instrument Co.,Ltd	Not Comply	One Local Agent with Seven Bidder	
12	Bistos Co.,Ltd	Not Comply	Bid Security Validity Period: February 18, 2024 Against February 27, 2024	
13	Joh. Acheils & Soehne GMBH	Not Comply	One Local Agent with Two Bidder	
14	Linnet Spot S.R.O.	Not Comply	One Local Agent with Three Bidder	
15	Beevejay Ltd	Not Comply	For item No 55,56&57 Payment Method: 100% LC @Sight Against 90% LC@Sight and the remaining 10% after installation.	
16	Media Biomedical Co.,Ltd	Not Comply	One Local Agent with Seven Bidder	
17	Nanjing Mindray Bio-Medical Electronics Co., Ltd	Not Comply	One Local Agent with Five Bidder	
18	Quayle Dental	Not Comply	Bid Security Validity Period: February 15, 2024 Against February 27, 2024	
19	Ningbo David Medical Devices Co., Ltd	Not Comply	One Local Agent with Five Bidder	
20	Niwe Sanayi Malzemelet	Not Comply	One Local Agent with Three Bidder	
21	Schiller AG	Not Comply	One Local Agent with Three Bidder	
22	Sharma Orthopedic Pvt.Ltd	Not Comply	Bid Security Period: 120 Days Against 148 Days	
23	Shenzhen Anke High-Tech Co., Ltd.	Not Comply	Payment Method: 100% LC@Sight and the remaining 20% after installation.	
24	Shenzhen Mindray Biomedical Electronics Co., Ltd	Not Comply	One Local Agent with Five Bidder	
25	Sonoscape Medical Corp	Not Comply	One Local Agent with Seven Bidder	
26	Qingdao Hater Biomedical Co., Ltd	Not Comply	Bid Security Validity Period: November 15, 2023 Against February 27, 2024	
27	Tolon Global Makina San. Tic.A.S	Not Comply	One Local Agent with Three Bidder	
28	Turkuaz Biyomedikal Tekn. Vazag. Vizm. San. Tic. Ltd. Sti	Not Comply	One Local Agent with Seven Bidder	
29	Tuttnaur Ltd	Not Comply	One Local Agent with Five Bidder	

አስታውሰው ይኒሩ ለመጠቀም በኋላ ትክክለኛ መሆኑን ያረጋግጡ  
 Make sure that this is the correct issue before use

OF for preliminary Evaluation, OF/TM/HO-02, 0, 21/01/2022





Document No: OF/TM/HO-15  
 Effective date : 21/01/2022  
 Company Name: የኢትዮጵያ ማዕከላዊ አጭሪ ኤጀንሲ  
 Ethiopian Pharmaceuticals Supply Agency  
 Title: Unsuccessful Bidder Under Financial Evaluation  
 Exchange Date: Oct 02, 2023  
 Rev. No.: 01  
 Page 1 of 1

Item No	Description	Unit	Qty	Bidder	Compliance	Reason of Non Compliance	Remark
2	MOBILE X-RAY MACHINE DIGITAL	Each	10	Jiangsu Saikang Co.,Ltd	Not Comply	Price Incompetee	
3	DIGITAL X-RAY MACHINE	Each	15	Jiangsu Saikang Co.,Ltd	Not Comply	Price Incompetee	
6	OXYGEN CONCENTRATOR	Each	50	Adonai Trading FZE	Not Comply	Price Incompetee	
11	STREAM STERILIZER	Each	30	Jiangsu Saikang Co.,Ltd	Not Comply	Price Incompetee	
20	STRETCHERS	Each	50	Adonai Trading FZE	Not Comply	Price Incompetee	
44	PEDIATRIC PATIENT BED WITH MATTRESS	Each	200	Adonai Trading FZE	Not Comply	Price Incompetee	
59	INFANT INCUBATOR	Each	50	Beeveejay Ltd	Not Comply	Price Incompetee	
67	LARGE BONE POWER DRILL WITH SAW SET	Set	10	Jiangsu Saikang Co.,Ltd	Not Comply	Price Incompetee	
68	SET - INSTRUMENT, RUSH NAIL	Set	10	Rise Global Gmbh	Not Comply	Price Incompetee	
72	SET - FIXATOR, EXTERNAL, LARGE (SYNTHEIS TYPE)	Set	10	Greens Surgicals Pvt.Ltd	Not Comply	Price Incompetee	
76	SET - INSTRUMENT, LARGE FRAGMENT LOCKING	Set	10	Greens Surgicals Pvt.Ltd	Not Comply	Price Incompetee	
80	CPAP MACHINE	Each	4	Adonai Trading FZE	Not Comply	Price Incompetee	

አስፈላጊ ማሳሰቢያ: በዚህ ሰነድ ላይ ያሉትን ጽሑፎች ማረጋገጥ ይገባል።  
 Make sure that this is the correct issue before use



Handwritten signatures and initials in blue ink.

OF for Financial Evaluation, OF/TM/HO-14 0, 21/01/2022