

To all participants on tender no.:- ICB/FW/EPSS6/SDG/ELISA/BLOOD BAG/MS/LR/28/23
Subject: Amendement on the bidding document and Bid Closing/opening date extension

A. Product Specification requirements Amendement

For item No. 8; -Blood transfusion set

- Soft and kink resistant, Sterile, Non toxic pyrogen free, PVC blood transfusion set
- Needle size 18G, can drop 20 drops/ml
- Standard blood transfusion set contain 170-200 micron filters
- Strong sleek and sharp spike, fitted with spike protector, without air vent
- Double drip clearly visible chamber facilitates visual access and rapid adjustment of fluid level
- Designed roller (thumb) controller offer accurate regulation of infusion rate
- Tube length not less than 1.3m

B. The following item is added in the above captioned tender reference number:-

ItemN	Item description	Unit	Quant	Quantity	Quantity	Rema
0,			ity	(2018	(2019	rk
			(2017	EC)	EC)	
			EC)			
	Blood transfusion set for pediatrics	each	40,000	80,000	120,000	
9	(IV blood transfusion set)					
	•Soft and kink resistant, Sterile, Non toxic pyrogen free, PVC				D	
	blood transfusion set					
	The second secon					

Proud to Serve ISO 9001 – 2015 Quality Certified AUTHORIZED ECONOMIC OPERATOR ID NO.: AEO/186/23 Please refer to our letter reference number when replying

Head Office
Address: Addls Ababa, Ethlopla
011-2763268 Director General
011-2763276 Public Relation Directorate
Hotline 8772 | P.O.Box. 21904
info@epss.gov.et | www.epss.gov.et



	Needle size 18G, can drop 20 drops/ml				
	•Standard blood transfusion set contain 170-200 micron filters			B	
	•Strong sleek and sharp spike, fitted with spike protector, without air vent			1	
	*Double drip clearly visible chamber facilitates visual access and rapid adjustment of fluid levelansfusion)	-			
	 Designed roller (thumb) controller offer accurate regulation of infusion rate Tube length not less than 150 cm 				
			à .		
				-	

C. The Bid closing and opening date and time is extended to April 11, 2024, Whereas other terms and conditions, evaluation criteria and requirements of the bidding document will remain the same.

NB: The price validity and bid bond validity need to be considered from extended bid

CC: Supplies agencial rector of the state of

- Deputy Director General(Inbound logistic)
- > Chemicals and Medical supplies procurement team
- > Archive

EPSS