

**To: All Participants of tender with Ref. No. ET-MOH-400012-GO-RFB**

**Subject: - Clarification on Tender with Ref. No. . ET-MOH-400012-GO-RFB**

Dear Sir/Madam;

We highly appreciate your interest to participate on the tender captioned above and reference is made to your kind request.

➤ **For Item No. 4 :- X-Ray machine:**

- The requested UPS with stabilizer is for the whole system.
- Please omit the phantom and be informed that the requested meter should be supplied with each X-Ray machine.
- The requested workstation is two with two monitor each and extra two monitors for PACS; Totally 6 Monitors are required.

➤ **For Item No. 5:- Anesthesia machine**

- Please kindly omit the word TEC-7 and for the rest proceed as per the bid document specification.

We hope the above points will clarify your request.

With Best Regards

CC:

➤ Medical Device Tender Management Team

➤ Archive

EPSS

  
Tizbeu Hailu Mogessie  
Pharmaceuticals & Medical  
Supplies Procurement Tender  
Management Directorate  
P/Director

